



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB127147

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	09/06/2023	11:40
2	CAL2	1	Water	NA	NA	20.7	7.00	09/06/2023	11:41
3	CAL3	1	Water	NA	NA	20.7	10.02	09/06/2023	11:43
4	ICV	1	Water	NA	NA	20.6	6.98	09/06/2023	11:45
5	CCV1	1	Water	NA	NA	20.6	2.02	09/06/2023	12:05
6	O4294-01	1	Water	NA	NA	21.2	5.60	09/06/2023	12:10
7	O4294-01DUP	1	Water	NA	NA	21.2	5.61	09/06/2023	12:11
8	CCV2	1	Water	NA	NA	20.6	12.02	09/06/2023	12:15

WORKLIST(Hardcopy Internal Chain)

VB 127167

WorkList Name : ph o4294 WorkList ID : 173324 Department : Wet-Chemistry Date : 09-06-2023 11:27:32

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4294-01	POOL-WATER	Water	pH	Cool 4 deg C	STER01	N11	08/31/2023	9040C

Date/Time 09/06/23 11:35
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 09/06/23 14:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]