


LB127153

245.1

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	(Std)
0.0	0.000	-0.012	-0.012	90	0.000	90					1
0.05	0.050	0.040	-0.010	344	0.0 %	344	0				-20
0.2	0.200	0.167	-0.033	971	0.0 %	971					-17
2.5	2.500	2.438	-0.062	12185	0.0 %	12185					-2
5.0	5.000	5.151	0.151	25579	0.0 %	25579					3
7.5	7.500	7.627	0.127	37800	0.0 %	37800					2
10.0	10.000	9.838	-0.162	48718	0.0 %	48718					-2

LB127153 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 06 Sep 2023 15:13:41

Sample ID	Extended ID	µ Abs.	Conc.	Std Cond Method	Units	Date	Type	Type
0.0 - 1	50	90	-	0.0000245.1	PPB	06 Sep 2023 15:31:37	S	Std
0.05 - 1	50.5	344	-	0.0500245.1	PPB	06 Sep 2023 15:33:54	S	Std
0.2 - 1	50.2	971	-	0.2000245.1	PPB	06 Sep 2023 15:41:11	S	Std
2.5 - 1	52.5	12185	-	2.5000245.1	PPB	06 Sep 2023 15:43:27	S	Std
5.0 - 1	55	25579	-	5.0000245.1	PPB	06 Sep 2023 15:45:44	S	Std
7.5 - 1	59.5	37800	-	7.5000245.1	PPB	06 Sep 2023 15:47:59	S	Std
10.0 - 1	518	48718	-	10.0000245.1	PPB	06 Sep 2023 15:52:50	S	Std
ICV48 - 1	ICV48	20266	4.0752	-245.1	PPB	06 Sep 2023 15:58:28	U	SMPL
ICB48 - 1	ICB48	105	-0.0085	-245.1	PPB	06 Sep 2023 16:00:43	U	SMPL
CCV51 - 1	CCV51	26019	5.2405	-245.1	PPB	06 Sep 2023 16:02:58	U	SMPL
CCB51 - 1	CCB51	105	-0.0085	-245.1	PPB	06 Sep 2023 16:05:13	U	SMPL
CRA - 1	CRA	325	0.0360	-245.1	PPB	06 Sep 2023 16:10:09	U	SMPL
HighStd - 1	HighStd	49334	9.9631	-245.1	PPB	06 Sep 2023 16:12:24	U	SMPL
ChkStd - 1	ChkStd	36080	7.2784	-245.1	PPB	06 Sep 2023 16:14:39	U	SMPL
PB155359BL - 1	PBW	-31	-0.0361	-245.1	PPB	06 Sep 2023 16:16:55	U	SMPL
PB155359BS - 1	LCSW	18150	3.6466	-245.1	PPB	06 Sep 2023 16:19:11	U	SMPL
O4013-01 - 1	EFFLUENT-DAY-1-4-CN	1727	0.3200	-245.1	PPB	06 Sep 2023 16:27:27	U	SMPL
O4013-01DUP - 1	EFFLUENT-DAY-1-4-CNDUP	1532	0.2805	-245.1	PPB	06 Sep 2023 16:35:12	U	SMPL
O4013-01MS - 1	EFFLUENT-DAY-1-4-CNMS	16665	3.3458	-245.1	PPB	06 Sep 2023 16:37:29	U	SMPL
O4013-01MSD - 1	EFFLUENT-DAY-1-4-CNMSD	16997	3.4130	-245.1	PPB	06 Sep 2023 16:44:48	U	SMPL
CCV52 - 1	CCV52	25255	5.0857	-245.1	PPB	06 Sep 2023 16:47:03	U	SMPL
CCB52 - 1	CCB52	157	0.0020	-245.1	PPB	06 Sep 2023 16:49:18	U	SMPL
O4013-01LX5 - 1		387	0.0486	-245.1	PPB	06 Sep 2023 16:55:14	U	SMPL
O4013-01A - 1		20126	4.0468	-245.1	PPB	06 Sep 2023 16:57:29	U	SMPL
CCV53 - 1	CCV53	25347	5.1044	-245.1	PPB	06 Sep 2023 16:59:45	U	SMPL
CCB53 - 1	CCB53	164	0.0034	-245.1	PPB	06 Sep 2023 17:02:00	U	SMPL