

LB127163

7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

58854

A= 0.0000e+000

B= 1.6882e-004

C= 5.8988e-002

Rho= 0.9998932

Accept=Accepted

Max. Conc: 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	b/p D
0.0	0.000	0.066	0.066	43	0.000	43	0				
0.2	0.200	0.179	-0.021	709	0.0 %	709					-1)
2.5	2.500	2.475	-0.025	14309	0.0 %	14309					-1)
5.0	5.000	4.916	-0.084	28770	0.0 %	28770					-2
7.5	7.500	7.570	0.070	44490	0.0 %	44490					1
10.0	10.000	9.995	-0.005	58854	0.0 %	58854					0

LB127163 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 07 Sep 2023 12:25:58

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	43	-	0.000	7470A	PPB	07 Sep 2023 12:39:21	S	Std
0.2 - 1	50-2	709	-	0.200	7470A	PPB	07 Sep 2023 12:41:38	S	Std
2.5 - 1	52.5	14309	-	2.500	7470A	PPB	07 Sep 2023 12:43:55	S	Std
5.0 - 1	55	28770	-	5.000	7470A	PPB	07 Sep 2023 12:46:12	S	Std
7.5 - 1	57.5	44490	-	7.500	7470A	PPB	07 Sep 2023 12:48:28	S	Std
10.0 - 1	510	58854	-	10.000	7470A	PPB	07 Sep 2023 12:55:52	S	Std
ICV49 - 1	ICV49	23754	4.0691	-	7470A	PPB	07 Sep 2023 12:58:38	U	SMPL
ICB49 - 1	ICB49	-18	0.0559	-	7470A	PPB	07 Sep 2023 13:00:53	U	SMPL
CCV54 - 1	CCV54	28246	4.8275	-	7470A	PPB	07 Sep 2023 13:03:07	U	SMPL
CCB54 - 1	CCB54	-38	0.0526	-	7470A	PPB	07 Sep 2023 13:05:22	U	SMPL
CRA - 1	CRA	663	0.1709	-	7470A	PPB	07 Sep 2023 13:07:37	U	SMPL
HighStd - 1	HighStd	58184	9.8816	-	7470A	PPB	07 Sep 2023 13:09:52	U	SMPL
ChkStd - 1	ChkStd	39902	6.7952	-	7470A	PPB	07 Sep 2023 13:12:08	U	SMPL
PB155382BL - 1	PBW	-23	0.0551	-	7470A	PPB	07 Sep 2023 13:14:23	U	SMPL
PB155382BS - 1	LCSW	19264	3.3111	-	7470A	PPB	07 Sep 2023 13:16:38	U	SMPL
O4256-11 - 1	Groundwater	-1	0.0588	-	7470A	PPB	07 Sep 2023 13:18:55	U	SMPL
O4273-01 - 1	1815	5584	1.0017	-	7470A	PPB	07 Sep 2023 13:21:11	U	SMPL
O4274-03 - 1	RB22282	1324	0.2825	-	7470A	PPB	07 Sep 2023 13:23:28	U	SMPL
O4275-01 - 1	60065	2080	0.4101	-	7470A	PPB	07 Sep 2023 13:25:44	U	SMPL
O4278-01 - 1	822-823	2329	0.4522	-	7470A	PPB	07 Sep 2023 13:27:59	U	SMPL
CCV55 - 1	CCV55	27991	4.7844	-	7470A	PPB	07 Sep 2023 13:32:49	U	SMPL
CCB55 - 1	CCB55	26	0.0634	-	7470A	PPB	07 Sep 2023 13:35:04	U	SMPL
O4290-01 - 1	402	554	0.1525	-	7470A	PPB	07 Sep 2023 13:37:19	U	SMPL
O4297-01 - 1	1728	1933	0.3853	-	7470A	PPB	07 Sep 2023 13:39:34	U	SMPL
O4313-02 - 1	COMP	1297	0.2779	-	7470A	PPB	07 Sep 2023 13:41:49	U	SMPL
O4313-02DUP - 1	COMPDUP	1225	0.2658	-	7470A	PPB	07 Sep 2023 13:44:05	U	SMPL
O4313-02MS - 1	COMPMS	12689	2.2011	-	7470A	PPB	07 Sep 2023 13:46:21	U	SMPL
O4313-02MSD - 1	COMPMSD	12843	2.2271	-	7470A	PPB	07 Sep 2023 13:48:37	U	SMPL
O4313-02LX5 - 1		36	0.0651	-	7470A	PPB	07 Sep 2023 13:50:55	U	SMPL
O4313-02A - 1		20242	3.4762	-	7470A	PPB	07 Sep 2023 13:55:57	U	SMPL
CCV56 - 1	CCV56	29446	5.0301	-	7470A	PPB	07 Sep 2023 14:03:14	U	SMPL
CCB56 - 1	CCB56	-21	0.0554	-	7470A	PPB	07 Sep 2023 14:05:29	U	SMPL