

LB127164


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

63955



A= 0.0000e+000

B= 1.5580e-004

C= -2.7708e-002

Rho= 0.9999347

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/1/1
0.0	0.000	-0.014	-0.014	85	0.000	85					-13
0.2	0.200	0.174	-0.026	1297	0.0 %	1297					0
2.5	2.500	2.498	-0.002	16211	0.0 %	16211					04
5.0	5.000	5.063	0.063	32675	0.0 %	32675					1
7.5	7.500	7.542	0.042	48588	0.0 %	48588					1
10.0	10.000	9.937	-0.063	63955	0.0 %	63955					1

LB127164 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 07 Sep 2023 14:16:05

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	85	-	0.0000	7471B	PPB	07 Sep 2023 14:21:00	S	Std
0.2 - 1	50.2	1297	-	0.2000	7471B	PPB	07 Sep 2023 14:23:17	S	Std
2.5 - 1	52.5	16211	-	2.5000	7471B	PPB	07 Sep 2023 14:25:33	S	Std
5.0 - 1	55	32675	-	5.0000	7471B	PPB	07 Sep 2023 14:27:50	S	Std
7.5 - 1	57.5	48588	-	7.5000	7471B	PPB	07 Sep 2023 14:30:06	S	Std
10.0 - 1	510	63955	-	10.0000	7471B	PPB	07 Sep 2023 14:32:22	S	Std
ICV50 - 1	ICV50	26514	4.1032	-	7471B	PPB	07 Sep 2023 14:38:23	U	SMPL
ICB50 - 1	ICB50	-32	-0.0327	-	7471B	PPB	07 Sep 2023 14:40:37	U	SMPL
CCV57 - 1	CCV57	32324	5.0084	-	7471B	PPB	07 Sep 2023 14:42:52	U	SMPL
CCB57 - 1	CCB57	-40	-0.0339	-	7471B	PPB	07 Sep 2023 14:45:07	U	SMPL
CRA - 1	CRA	1243	0.1660	-	7471B	PPB	07 Sep 2023 14:47:23	U	SMPL
HighStd - 1	HighStd	63122	9.8068	-	7471B	PPB	07 Sep 2023 14:49:39	U	SMPL
ChkStd - 1	ChkStd	45319	7.0331	-	7471B	PPB	07 Sep 2023 14:51:54	U	SMPL
PB155383BL - 1	PBS	-90	-0.0417	-	7471B	PPB	07 Sep 2023 14:54:11	U	SMPL
PB155383BS - 1	LCSS	24912	3.8536	-	7471B	PPB	07 Sep 2023 15:08:33	U	SMPL
O4278-02 - 1	824-829	4225	0.6306	-	7471B	PPB	07 Sep 2023 15:10:49	U	SMPL
O4283-06 - 1	C-1	2770	0.4039	-	7471B	PPB	07 Sep 2023 15:13:05	U	SMPL
O4283-06DUP - 1	C-1DUP	3024	0.4434	-	7471B	PPB	07 Sep 2023 15:15:21	U	SMPL
O4283-06MS - 1	C-1MS	28761	4.4533	-	7471B	PPB	07 Sep 2023 15:17:38	U	SMPL
O4283-06MSD - 1	C-1MSD	28309	4.3829	-	7471B	PPB	07 Sep 2023 15:19:54	U	SMPL
CCV58 - 1	CCV58	32412	5.0221	-	7471B	PPB	07 Sep 2023 15:22:08	U	SMPL
CCB58 - 1	CCB58	-34	-0.0330	-	7471B	PPB	07 Sep 2023 15:24:24	U	SMPL
O4283-07 - 1	C-2	3893	0.5788	-	7471B	PPB	07 Sep 2023 15:26:39	U	SMPL
O4283-08 - 1	C-3	5283	0.7954	-	7471B	PPB	07 Sep 2023 15:28:53	U	SMPL
O4283-06LX5 - 1		550	0.0580	-	7471B	PPB	07 Sep 2023 15:34:10	U	SMPL
O4283-06A - 1		28816	4.4619	-	7471B	PPB	07 Sep 2023 15:36:25	U	SMPL
CCV59 - 1	CCV59	32581	5.0485	-	7471B	PPB	07 Sep 2023 15:38:40	U	SMPL
CCB59 - 1	CCB59	-24	-0.0314	-	7471B	PPB	07 Sep 2023 15:40:56	U	SMPL