



PERCENT SOLID

Supervisor: Iwona

Analyst: JIGNESH

Date: 9/8/2023

OVENTEMP IN Celsius(°C): 107

Time IN: 17:00

In Date: 09/07/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:00

Out Date: 09/08/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB127170

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O4310-01	SCF	1	1.15	8.50	9.65	9.3	95.9	
O4322-01	COMP-1	2	1.00	1.00	2.00	2.00	100.0	OIL SAMPLE
O4333-01	WC-1	3	1.19	8.67	9.86	8.16	80.4	
O4333-02	WC-1-EPH	4	1.19	8.67	9.86	8.16	80.4	
O4334-01	359	5	1.13	8.76	9.89	9.09	90.9	
O4334-02	359	6	1.13	8.68	9.81	9.07	91.5	
O4335-01	AREA-1	21	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
O4335-02	AREA-1	22	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
O4335-03	AREA-2	23	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
O4335-04	AREA-2	24	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
O4336-01	TRENCH-A	7	1.19	8.62	9.81	8.64	86.4	
O4336-02	TRENCH-A-E2	8	1.19	8.73	9.92	8.7	86.0	
O4336-03	TRENCH-B	9	1.18	8.66	9.84	8.76	87.5	
O4336-04	TRENCH-B-E2	10	1.18	8.48	9.66	8.54	86.8	
O4336-05	DUCT-BANK-A	11	1.14	8.63	9.77	8.64	86.9	
O4336-06	DUCT-BANK-A-E2	12	1.16	8.83	9.99	8.62	84.5	
O4336-07	DUCT-BANK-B	13	1.19	8.71	9.9	8.77	87.0	
O4336-08	DUCT-BANK-B-E2	14	1.13	8.62	9.75	8.7	87.8	
O4338-01	TP4	15	1.18	8.68	9.86	9.06	90.8	
O4338-02	TP4-EPH	16	1.18	8.68	9.86	9.06	90.8	
O4338-04	TP3	17	1.19	8.67	9.86	9.12	91.5	
O4338-05	TP3-EPH	18	1.19	8.67	9.86	9.12	91.5	
O4338-07	TP2	19	1.11	8.88	9.99	8.79	86.5	
O4338-08	TP2-EPH	20	1.11	8.88	9.99	8.79	86.5	
O4340-01	40277	25	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O4340-02	40278	26	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$