



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB127177

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	09/07/2023	10:20
2	O4297-01		1	12	100.00	>212.0	>212.0	09/07/2023	10:50
3	O4297-01DUP		1	12	100.00	>212.0	>212.0	09/07/2023	11:20
4	O4297-02		1	14	100.00	>212.0	>212.0	09/07/2023	11:50
5	O4297-03		1	12	100.00	>212.0	>212.0	09/07/2023	12:20
6	O4297-04		1	13	100.00	>212.0	>212.0	09/07/2023	12:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

66127177

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp-9-6

WorkList ID : 173303

Department : Wet-Chemistry

Date : 09-06-2023 07:53:15

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4297-01	1728	Water	Flash Point	Cool 4 deg C	PSEG03	N11	09/05/2023	1010B
O4297-02	1683	Water	Flash Point	Cool 4 deg C	PSEG03	N11	09/05/2023	1010B
O4297-03	713-G-DIP	Water	Flash Point	Cool 4 deg C	PSEG03	N11	09/05/2023	1010B
O4297-04	1731-DRIP	Water	Flash Point	Cool 4 deg C	PSEG03	N11	09/05/2023	1010B

Date/Time 09/07/2023 10.10
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: RM (121)

Date/Time 09/07/2023 13.30
 Raw Sample Received by: RM (DL)
 Raw Sample Relinquished by: RM CWC