

LB127182


7470A

INSTRUMENT ID: EYJ

Linear

μ Abs.:

5472E



A= 0.0000e+000

B= 1.8180e-004

C= -5.0741e-003

Rho= 0.9999235

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%ID
0.0	0.000	0.007	0.007	67	0.000	67	0				1
0.2	0.200	0.204	0.004	1152	0.0 %	1152					2
2.5	2.500	2.443	-0.057	13465	0.0 %	13465					-2
5.0	5.000	5.029	0.029	27692	0.0 %	27692					1
7.5	7.500	7.572	0.072	41678	0.0 %	41678					1
10.0	10.000	9.944	-0.056	54726	0.0 %	54726					-1

LB127182 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 08 Sep 2023 09:02:21

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	67	-	0.0000	7470A	PPB	08 Sep 2023 09:07:39	S	Std
0.2 - 1	50.2	1152	-	0.2000	7470A	PPB	08 Sep 2023 09:09:55	S	Std
2.5 - 1	52.5	13465	-	2.5000	7470A	PPB	08 Sep 2023 09:12:12	S	Std
5.0 - 1	55	27692	-	5.0000	7470A	PPB	08 Sep 2023 09:14:29	S	Std
7.5 - 1	57.5	41678	-	7.5000	7470A	PPB	08 Sep 2023 09:16:48	S	Std
10.0 - 1	510	54726	-	10.0000	7470A	PPB	08 Sep 2023 09:19:08	S	Std
ICV51 - 1	ICV51	23222	4.2167	-7470A	PPB	08 Sep 2023 09:21:59	U	SMPL	
ICB51 - 1	ICB51	-26	-0.0098	-7470A	PPB	08 Sep 2023 09:24:14	U	SMPL	
CCV60 - 1	CCV60	26012	4.7240	-7470A	PPB	08 Sep 2023 09:26:29	U	SMPL	
CCB60 - 1	CCB60	-22	-0.0091	-7470A	PPB	08 Sep 2023 09:28:44	U	SMPL	
CRA - 1	CRA	1100	0.1949	-7470A	PPB	08 Sep 2023 09:30:58	U	SMPL	
HighStd - 1	HighStd	54656	9.9315	-7470A	PPB	08 Sep 2023 09:33:13	U	SMPL	
ChkStd - 1	ChkStd	39291	7.1381	-7470A	PPB	08 Sep 2023 09:35:28	U	SMPL	
PB155398BL - 1	PBW	-58	-0.0156	-7470A	PPB	08 Sep 2023 09:37:43	U	SMPL	
PB155398BS - 1	LCSW	22474	4.0807	-7470A	PPB	08 Sep 2023 09:39:59	U	SMPL	
O4256-02 - 1	MW-1	-19	-0.0085	-7470A	PPB	08 Sep 2023 09:42:15	U	SMPL	
O4256-02DUP - 1	MW-1DUP	27	-0.0002	-7470A	PPB	08 Sep 2023 09:44:31	U	SMPL	
O4256-02MS - 1	MW-1MS	22023	3.9987	-7470A	PPB	08 Sep 2023 09:46:48	U	SMPL	
O4256-02MSD - 1	MW-1MSD	19839	3.6017	-7470A	PPB	08 Sep 2023 09:49:04	U	SMPL	
O4256-04 - 1	MW-2	-10	-0.0069	-7470A	PPB	08 Sep 2023 09:51:19	U	SMPL	
CCV61 - 1	CCV61	26233	4.7641	-7470A	PPB	08 Sep 2023 09:53:34	U	SMPL	
CCB61 - 1	CCB61	-36	-0.0116	-7470A	PPB	08 Sep 2023 09:55:49	U	SMPL	
O4256-06 - 1	MW-3	26	-0.0003	-7470A	PPB	08 Sep 2023 09:58:04	U	SMPL	
O4256-08 - 1	MW-4-6	27	-0.0002	-7470A	PPB	08 Sep 2023 10:00:19	U	SMPL	
O4256-10 - 1	MW-8R-4R-9	26	-0.0003	-7470A	PPB	08 Sep 2023 10:02:35	U	SMPL	
O4258-01 - 1	PAINT-CHIPS-COMP	82	0.0098	-7470A	PPB	08 Sep 2023 10:04:50	U	SMPL	
O4274-02 - 1	RT2683	101	0.0133	-7470A	PPB	08 Sep 2023 10:07:06	U	SMPL	
PB155399BL - 1	PBW	38	0.0018	-7470A	PPB	08 Sep 2023 10:09:22	U	SMPL	
PB155399BS - 1	LCSW	21804	3.9589	-7470A	PPB	08 Sep 2023 10:11:38	U	SMPL	
O4291-01 - 1	SOIL-CEMENT-COMP	-1	-0.0053	-7470A	PPB	08 Sep 2023 10:13:55	U	SMPL	
O4299-02 - 1	SB-01-COMP	38	0.0018	-7470A	PPB	08 Sep 2023 10:16:11	U	SMPL	
O4299-04 - 1	SB-02-COMP	75	0.0086	-7470A	PPB	08 Sep 2023 10:18:26	U	SMPL	
CCV62 - 1	CCV62	26106	4.7410	-7470A	PPB	08 Sep 2023 10:20:42	U	SMPL	
CCB62 - 1	CCB62	-25	-0.0096	-7470A	PPB	08 Sep 2023 10:22:57	U	SMPL	
O4299-06 - 1	SB-03-COMP	514	0.0884	-7470A	PPB	08 Sep 2023 10:25:12	U	SMPL	
O4299-08 - 1	SB-04-COMP	100	0.0131	-7470A	PPB	08 Sep 2023 10:27:27	U	SMPL	
O4305-03 - 1	TP-9-5-23-EPH	49	0.0038	-7470A	PPB	08 Sep 2023 10:29:42	U	SMPL	
O4305-03DUP - 1	TP-9-5-23-EPHDUP	56	0.0051	-7470A	PPB	08 Sep 2023 10:31:58	U	SMPL	
O4305-03MS - 1	TP-9-5-23-EPHMS	20706	3.7593	-7470A	PPB	08 Sep 2023 10:34:13	U	SMPL	
O4305-03MSD - 1	TP-9-5-23-EPHMSD	21783	3.9551	-7470A	PPB	08 Sep 2023 10:36:29	U	SMPL	
PB155300TB - 1	PB155300TB	-52	-0.0145	-7470A	PPB	08 Sep 2023 10:38:45	U	SMPL	
PB155351TB - 1	PB155351TB	13	-0.0027	-7470A	PPB	08 Sep 2023 10:41:02	U	SMPL	
CCV63 - 1	CCV63	26737	4.8558	-7470A	PPB	08 Sep 2023 10:43:19	U	SMPL	
CCB63 - 1	CCB63	-22	-0.0091	-7470A	PPB	08 Sep 2023 10:45:34	U	SMPL	
O4256-02LX5 - 1		-6	-0.0062	-7470A	PPB	08 Sep 2023 10:47:50	U	SMPL	
O4256-02A - 1		23428	4.2542	-7470A	PPB	08 Sep 2023 10:50:05	U	SMPL	
O4305-03LX5 - 1		-37	-0.0118	-7470A	PPB	08 Sep 2023 10:52:20	U	SMPL	
O4305-03A - 1		24002	4.3585	-7470A	PPB	08 Sep 2023 10:54:35	U	SMPL	
CCV64 - 1	CCV64	27730	5.0363	-7470A	PPB	08 Sep 2023 10:56:51	U	SMPL	
CCB64 - 1	CCB64	-39	-0.0122	-7470A	PPB	08 Sep 2023 10:59:06	U	SMPL	