



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB127188

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	09/08/2023	09:05
2	CAL2	1	Water	NA	NA	20.6	7.00	09/08/2023	09:06
3	CAL3	1	Water	NA	NA	20.6	10.02	09/08/2023	09:09
4	ICV	1	Water	NA	NA	20.7	7.01	09/08/2023	09:10
5	CCV1	1	Water	NA	NA	20.7	2.01	09/08/2023	09:15
6	O4316-01	1	Water	NA	NA	20.2	7.78	09/08/2023	09:20
7	O4337-01	1	Water	NA	NA	20.1	6.11	09/08/2023	09:25
8	O4341-01	1	Water	NA	NA	21.5	7.62	09/08/2023	09:27
9	O4341-02	1	Water	NA	NA	22.0	5.69	09/08/2023	09:34
10	O4343-01	1	Water	NA	NA	20.3	7.27	09/08/2023	09:40
11	O4343-02	1	Water	NA	NA	20.2	7.37	09/08/2023	09:45
12	O4343-02DUP	1	Water	NA	NA	20.3	7.38	09/08/2023	09:46
13	CCV2	1	Water	NA	NA	20.7	12.03	09/08/2023	09:50

WORKLIST(Hardcopy Internal Chain)

127188

WorkList Name : PH 03778

WorkList ID : 173385

Department : Wet-Chemistry

Date : 09-08-2023 07:41:03

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4316-01	A DRAIN-WATER-TANK-1	Water	pH	Cool 4 deg C	MAJO01	N11	09/06/2023	9040C
O4337-01	A DRAIN-WATER-TANK-1	Water	pH	Cool 4 deg C	MAJO01	N11	09/07/2023	9040C
O4341-01	A AUD-1046-TOTE	Water	pH	Cool 4 deg C	PSEG03	N11	09/07/2023	9040C
O4341-02	A AUD-1047	Water	pH	Cool 4 deg C	PSEG03	N11	09/07/2023	9040C
O4343-01	I SAMPLE-1-14TH-STREET	Water	pH	Cool 4 deg C	SOUN03	N11	09/06/2023	9040C
O4343-02	I SAMPLE-2-23RD-STREET	Water	pH	Cool 4 deg C	SOUN03	N11	09/06/2023	9040C

Date/Time 09/08/23 09:00

Raw Sample Received by: 12100

Raw Sample Relinquished by: 12100

Date/Time 09/08/23 12:00
Raw Sample Received by: 12100
Raw Sample Relinquished by: 12100