

LB127193


7471B

INSTRUMENT ID: EV1

Linear

 μ Abs.:

64309



A= 0.0000e+000

B= 1.5599e-004

C= -3.7534e-002

Rho= 0.9998771

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.032	-0.032	36	0.000	36					
0.2	0.200	0.167	-0.033	1314	0.0 %	1314					-17
2.5	2.500	2.547	0.047	16566	0.0 %	16566					2
5.0	5.000	5.100	0.100	32933	0.0 %	32933					2
7.5	7.500	7.424	-0.076	47830	0.0 %	47830					-1
10.0	10.000	9.994	-0.006	64309	0.0 %	64309					0

LB127193 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 08 Sep 2023 11:36:01

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	36	-	0.0000	7471B	PPB	08 Sep 2023 11:47:39	S	Std
0.2 - 1	50.2	1314	-	0.2000	7471B	PPB	08 Sep 2023 11:49:56	S	Std
2.5 - 1	52.5	16566	-	2.5000	7471B	PPB	08 Sep 2023 11:52:13	S	Std
5.0 - 1	55	32933	-	5.0000	7471B	PPB	08 Sep 2023 11:54:29	S	Std
7.5 - 1	57.5	47830	-	7.5000	7471B	PPB	08 Sep 2023 11:56:46	S	Std
10.0 - 1	51.8	64309	-	10.0000	7471B	PPB	08 Sep 2023 12:01:48	S	Std
ICV52 - 1	ICV52	25985	4.0160	-	7471B	PPB	08 Sep 2023 12:04:35	U	SMPL
ICB52 - 1	ICB52	-53	-0.0458	-	7471B	PPB	08 Sep 2023 12:06:50	U	SMPL
CCV65 - 1	CCV65	32609	5.0493	-	7471B	PPB	08 Sep 2023 12:13:23	U	SMPL
CCB65 - 1	CCB65	-54	-0.0460	-	7471B	PPB	08 Sep 2023 12:15:37	U	SMPL
CRA - 1	CRA	1296	0.1646	-	7471B	PPB	08 Sep 2023 12:17:53	U	SMPL
HighStd - 1	HighStd	64668	10.0503	-	7471B	PPB	08 Sep 2023 12:20:07	U	SMPL
ChkStd - 1	ChkStd	44557	6.9131	-	7471B	PPB	08 Sep 2023 12:22:22	U	SMPL
PB155400BL - 1	PBS	-59	-0.0467	-	7471B	PPB	08 Sep 2023 12:24:39	U	SMPL
PB155400BS - 1	LCSS	23011	3.5521	-	7471B	PPB	08 Sep 2023 12:26:56	U	SMPL
O4256-01 - 1	MW-1	756	0.0804	-	7471B	PPB	08 Sep 2023 12:29:12	U	SMPL
O4256-01DUP - 1	MW-1DUP	701	0.0718	-	7471B	PPB	08 Sep 2023 12:31:28	U	SMPL
O4256-01MS - 1	MW-1MS	26976	4.1706	-	7471B	PPB	08 Sep 2023 12:33:45	U	SMPL
O4256-01MSD - 1	MW-1MSD	28180	4.3584	-	7471B	PPB	08 Sep 2023 12:36:01	U	SMPL
O4256-03 - 1	MW-2	337	0.0150	-	7471B	PPB	08 Sep 2023 12:38:17	U	SMPL
CCV66 - 1	CCV66	33729	5.2240	-	7471B	PPB	08 Sep 2023 12:40:33	U	SMPL
CCB66 - 1	CCB66	-61	-0.0471	-	7471B	PPB	08 Sep 2023 12:42:48	U	SMPL
O4256-05 - 1	MW-3	5305	0.7900	-	7471B	PPB	08 Sep 2023 12:45:04	U	SMPL
O4256-07 - 1	MW-4-6	465	0.0350	-	7471B	PPB	08 Sep 2023 12:47:19	U	SMPL
O4256-09 - 1	MW-8R-4R-9	875	0.0990	-	7471B	PPB	08 Sep 2023 12:49:35	U	SMPL
O4274-01 - 1	RT2683	1040	0.1247	-	7471B	PPB	08 Sep 2023 12:51:50	U	SMPL
O4293-01 - 1	SOIL	3820	0.5584	-	7471B	PPB	08 Sep 2023 12:54:05	U	SMPL
O4304-04 - 1	WC-1	16866	2.5935	-	7471B	PPB	08 Sep 2023 12:56:22	U	SMPL
O4335-01 - 1	AREA-1	246	0.0008	-	7471B	PPB	08 Sep 2023 12:58:38	U	SMPL
O4335-03 - 1	AREA-2	441	0.0313	-	7471B	PPB	08 Sep 2023 13:00:55	U	SMPL
O4338-01 - 1	TP4	3199	0.4615	-	7471B	PPB	08 Sep 2023 13:03:12	U	SMPL
O4338-04 - 1	TP3	3954	0.5793	-	7471B	PPB	08 Sep 2023 13:05:28	U	SMPL
CCV67 - 1	CCV67	34286	5.3109	-	7471B	PPB	08 Sep 2023 13:07:43	U	SMPL
CCB67 - 1	CCB67	-38	-0.0435	-	7471B	PPB	08 Sep 2023 13:09:58	U	SMPL
O4338-07 - 1	TP2	3103	0.4465	-	7471B	PPB	08 Sep 2023 13:12:14	U	SMPL
PB155401BL - 1	PBS	-2	-0.0378	-	7471B	PPB	08 Sep 2023 13:14:29	U	SMPL
PB155401BS - 1	LCSS	24680	3.8124	-	7471B	PPB	08 Sep 2023 13:16:44	U	SMPL
O4305-01 - 1	TP-9-5-23-EPH	6808	1.0245	-	7471B	PPB	08 Sep 2023 13:18:59	U	SMPL
O4305-01DUP - 1	TP-9-5-23-EPHDUP	6688	1.0058	-	7471B	PPB	08 Sep 2023 13:23:59	U	SMPL
O4305-01MS - 1	TP-9-5-23-EPHMS	33653	5.2121	-	7471B	PPB	08 Sep 2023 13:26:15	U	SMPL
O4305-01MSD - 1	TP-9-5-23-EPHMSD	34218	5.3003	-	7471B	PPB	08 Sep 2023 13:28:31	U	SMPL
O4310-01 - 1	SCF	13655	2.0926	-	7471B	PPB	08 Sep 2023 13:30:48	U	SMPL
O4312-01 - 1	RT4213	1934	0.2642	-	7471B	PPB	08 Sep 2023 13:33:06	U	SMPL
O4314-01 - 1	VNJ-201	438	0.0308	-	7471B	PPB	08 Sep 2023 13:35:21	U	SMPL
CCV68 - 1	CCV68	33495	5.1875	-	7471B	PPB	08 Sep 2023 13:37:37	U	SMPL
CCB68 - 1	CCB68	1	-0.0374	-	7471B	PPB	08 Sep 2023 13:39:52	U	SMPL
O4321-01 - 1	RT2345	1424	0.1846	-	7471B	PPB	08 Sep 2023 13:42:07	U	SMPL
O4321-03 - 1	RT4296	17094	2.6290	-	7471B	PPB	08 Sep 2023 13:44:22	U	SMPL
O4333-01 - 1	WC-1	4666	0.6903	-	7471B	PPB	08 Sep 2023 13:46:38	U	SMPL
O4334-02 - 1	359	6697	1.0072	-	7471B	PPB	08 Sep 2023 13:48:53	U	SMPL
O4336-01 - 1	TRENCH-A	10041	1.5288	-	7471B	PPB	08 Sep 2023 13:51:09	U	SMPL
O4336-03 - 1	TRENCH-B	6739	1.0137	-	7471B	PPB	08 Sep 2023 13:53:26	U	SMPL
O4336-05 - 1	DUCT-BANK-A	12780	1.9561	-	7471B	PPB	08 Sep 2023 13:55:42	U	SMPL
O4336-07 - 1	DUCT-BANK-B	8858	1.3443	-	7471B	PPB	08 Sep 2023 13:57:59	U	SMPL
CCV69 - 1	CCV69	33369	5.1678	-	7471B	PPB	08 Sep 2023 14:00:19	U	SMPL
CCB69 - 1	CCB69	-7	-0.0386	-	7471B	PPB	08 Sep 2023 14:02:34	U	SMPL
O4256-01LX5 - 1		93	-0.0230	-	7471B	PPB	08 Sep 2023 14:04:50	U	SMPL
O4256-01A - 1		27804	4.2997	-	7471B	PPB	08 Sep 2023 14:07:05	U	SMPL
O4305-01LX5 - 1		1264	0.1596	-	7471B	PPB	08 Sep 2023 14:09:20	U	SMPL
O4305-01A - 1		33901	5.2508	-	7471B	PPB	08 Sep 2023 14:11:35	U	SMPL
CCV70 - 1	CCV70	33329	5.1616	-	7471B	PPB	08 Sep 2023 14:13:55	U	SMPL
CCB70 - 1	CCB70	-28	-0.0419	-	7471B	PPB	08 Sep 2023 14:16:11	U	SMPL