


LB127200

SFAM01.1

INSTRUMENT ID: CY1



Linear

A= 0.0000e+000

B= 1.8182e-004 slope

C= -3.3816e-002 y-intercept

Rho= 0.9999421

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.006	-0.006	151	0.000	151					=
0.05	0.050					501	0				-))
0.20	0.200	0.178	-0.022	1166	0.0 %	1166					0
2.50	2.500	2.499	-0.001	13932	0.0 %	13932					1
5.00	5.000	5.025	0.025	27824	0.0 %	27824					1
7.50	7.500	7.566	0.066	41799	0.0 %	41799					1
10.0	10.000	9.937	-0.063	54840	0.0 %	54840					-)

LB127200
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	151	-		0 SFAM01.1	PPB	9/8/2023 15:23	Std
	0.2 S01	1166	-		0.2 SFAM01.1	PPB	9/8/2023 15:26	Std
	2.5 S02	13932	-		2.5 SFAM01.1	PPB	9/8/2023 15:28	Std
	5 S03	27824	-		5 SFAM01.1	PPB	9/8/2023 15:30	Std
	7.5 S04	41799	-		7.5 SFAM01.1	PPB	9/8/2023 15:32	Std
	10 S05	54840	-		10 SFAM01.1	PPB	9/8/2023 15:40	Std
ICV054	ICV054	21116	3.8056 -		SFAM01.1	PPB	9/8/2023 15:43	SMPL
ICB054	ICB054	81	-0.0191 -		SFAM01.1	PPB	9/8/2023 15:45	SMPL
CCV071	CCV071	27518	4.9696 -		SFAM01.1	PPB	9/8/2023 15:47	SMPL
CCB071	CCB071	88	-0.0178 -		SFAM01.1	PPB	9/8/2023 15:49	SMPL
PB155428BL	PBW428	268	0.0149 -		SFAM01.1	PPB	9/8/2023 15:51	SMPL
O4114-12	MEZYN8	94	-0.0167 -		SFAM01.1	PPB	9/8/2023 15:54	SMPL
O4114-22	MEZYN9	140	-0.0084 -		SFAM01.1	PPB	9/8/2023 15:56	SMPL
PB155429BL	PBW429	286	0.0182 -		SFAM01.1	PPB	9/8/2023 15:58	SMPL
O4155-01	MCOAG8	508	0.0585 -		SFAM01.1	PPB	9/8/2023 16:01	SMPL
O4155-02	MCOAG8D	468	0.0513 -		SFAM01.1	PPB	9/8/2023 16:03	SMPL
O4155-03	MCOAG8S	5639	0.9915 -		SFAM01.1	PPB	9/8/2023 16:05	SMPL
PB155430BL	PBW430	233	0.0085 -		SFAM01.1	PPB	9/8/2023 16:07	SMPL
O4156-01	MCOAG9	270	0.0153 -		SFAM01.1	PPB	9/8/2023 16:10	SMPL
O4156-02	MCOAG9D	196	0.0018 -		SFAM01.1	PPB	9/8/2023 16:12	SMPL
O4156-03	MCOAG9S	6005	1.058 -		SFAM01.1	PPB	9/8/2023 16:14	SMPL
CCV072	CCV072	26937	4.864 -		SFAM01.1	PPB	9/8/2023 16:16	SMPL
CCB072	CCB072	99	-0.0158 -		SFAM01.1	PPB	9/8/2023 16:19	SMPL