



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB127204

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	09/08/2023	11:10
2	O4299-09		1	12	100.00	>212.0	>212.0	09/08/2023	11:40
3	O4299-09DUP		1	14	100.00	>212.0	>212.0	09/08/2023	12:10
4	O4299-10		1	12	100.00	>212.0	>212.0	09/08/2023	12:40
5	O4341-01		1	13	100.00	>212.0	>212.0	09/08/2023	13:10
6	O4341-02		1	14	100.00	>212.0	>212.0	09/08/2023	13:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

26127204

WorkList Name : FP W-9-7

WorkList ID : 173381

Department : Wet-Chemistry

Date : 09-07-2023 18:12:35

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4299-09 L	MW-01	Water	Flash Point	Cool 4 deg C	LIRO01	N11	09/05/2023	1010B
O4299-10 L	TWP-01	Water	Flash Point	Cool 4 deg C	LIRO01	N11	09/05/2023	1010B
O4341-01 L	AUD-1046-TOTE	Water	Flash Point	Cool 4 deg C	PSEG03	N11	09/07/2023	1010B
O4341-02 L	AUD-1047	Water	Flash Point	Cool 4 deg C	PSEG03	N11	09/07/2023	1010B

Date/Time 09/08/2023 11:00
 Raw Sample Received by: RM awc
 Raw Sample Relinquished by: RM awc

Date/Time 09/08/2023 14:05
 Raw Sample Received by: RM awc
 Raw Sample Relinquished by: RM awc