



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB127212

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	09/08/2023	15:10
2	CAL2	1	Water	NA	NA	20.6	7.00	09/08/2023	15:11
3	CAL3	1	Water	NA	NA	20.6	10.02	09/08/2023	15:12
4	ICV	1	Water	NA	NA	20.6	6.98	09/08/2023	15:15
5	CCV1	1	Water	NA	NA	20.7	2.01	09/08/2023	15:20
6	O4338-03	1	Solid	20.02	20	22.1	7.02	09/08/2023	15:22
7	O4338-03DUP	1	Solid	20.01	20	22.2	7.03	09/08/2023	15:23
8	O4338-06	1	Solid	20.03	20	23.6	8.20	09/08/2023	15:25
9	O4338-09	1	Solid	20.05	20	23.6	7.97	09/08/2023	15:30
10	CCV2	1	Water	NA	NA	20.7	12.02	09/08/2023	15:35

127212

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity 04338

WorkList ID : 173424

Department : Wet-Chemistry

Date : 09-08-2023 14:50:33

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4338-03	TP4	Solid	Corrosivity	Cool 4 deg C	PSEG03	N21	09/07/2023	9045D
O4338-06	TP3	Solid	Corrosivity	Cool 4 deg C	PSEG03	N21	09/07/2023	9045D
O4338-09	TP2	Solid	Corrosivity	Cool 4 deg C	PSEG03	N21	09/07/2023	9045D

Date/Time 09/08/23 15:00

Raw Sample Received by: 10 (WCC)

Raw Sample Relinquished by: JTC (sm)

Date/Time 09/08/23 17:00

Raw Sample Received by: JTC (sm)

Raw Sample Relinquished by: 10 (WCC)