



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB127223

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.7	09/11/2023	09:20
2	O4343-01		1	13	100.00	>212.0	>212.0	09/11/2023	09:50
3	O4343-01DUP		1	12	100.00	>212.0	>212.0	09/11/2023	10:20
4	O4343-02		1	14	100.00	>212.0	>212.0	09/11/2023	10:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

(6127223)

WorkList Name : FP-9-11

WorkList ID : 173458

Department : Wet-Chemistry

Date : 09-11-2023 08:29:25

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4343-01	SAMPLE-1-14TH-STREET	Water	Flash Point	Cool 4 deg C	SOUN03	N11	09/06/2023	1010B
O4343-02	SAMPLE-2-23RD-STREET	Water	Flash Point	Cool 4 deg C	SOUN03	N11	09/06/2023	1010B

Date/Time 09/11/2023 09:10
 Raw Sample Received by: R14
 Raw Sample Relinquished by: JDC

Date/Time 09/11/2023 11:30
 Raw Sample Received by: JDC
 Raw Sample Relinquished by: R14