

LB127225


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

6076E



A= 0.0000e+000

B= 1.6890e-004 slope

C= -2.6676e-001 intercept

Rho= 0.9999317

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.007	0.007	1622	0.000	1622					
0.05	0.050					501	0				
0.20	0.200	0.213	0.013	2843	0.0 %	2843					
2.50	2.500	2.432	-0.068	15981	0.0 %	15981					
5.00	5.000	5.076	0.076	31629	0.0 %	31629					
7.50	7.500	7.474	-0.026	45832	0.0 %	45832					
10.0	10.000	9.997	-0.003	60766	0.0 %	60766					

0.00
=7
-3
2
0
0

LB127225

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	1622	-		0 SFAM01.1	PPB	9/11/2023 13:21	Std
	0.2 S01	2843	-		0.2 SFAM01.1	PPB	9/11/2023 13:24	Std
	2.5 S02	15981	-		2.5 SFAM01.1	PPB	9/11/2023 13:26	Std
	5 S03	31629	-		5 SFAM01.1	PPB	9/11/2023 13:28	Std
	7.5 S04	45832	-		7.5 SFAM01.1	PPB	9/11/2023 13:30	Std
	10 S05	60766	-		10 SFAM01.1	PPB	9/11/2023 13:35	Std
ICV056	ICV056	26688	4.241 -		SFAM01.1	PPB	9/11/2023 13:38	SMPL
ICB056	ICB056	1567	-0.0021 -		SFAM01.1	PPB	9/11/2023 13:40	SMPL
CCV075	CCV075	31868	5.1159 -		SFAM01.1	PPB	9/11/2023 13:42	SMPL
CCB075	CCB075	1546	-0.0056 -		SFAM01.1	PPB	9/11/2023 13:45	SMPL
PB155457BL	PBS457	1566	-0.0023 -		SFAM01.1	PPB	9/11/2023 13:49	SMPL
O4114-01	MEZYH4	1671	0.0155 -		SFAM01.1	PPB	9/11/2023 13:52	SMPL
O4114-02	MEZYH5	4771	0.5391 -		SFAM01.1	PPB	9/11/2023 13:54	SMPL
O4114-03	MEZYH5D	4515	0.4958 -		SFAM01.1	PPB	9/11/2023 13:56	SMPL
O4114-04	MEZYH5S	16540	2.5269 -		SFAM01.1	PPB	9/11/2023 13:59	SMPL
O4114-05	MEZYH6	7750	1.0422 -		SFAM01.1	PPB	9/11/2023 14:01	SMPL
O4114-06	MEZYL1	25	-0.2625 -		SFAM01.1	PPB	9/11/2023 14:03	SMPL
O4114-07	MEZYL2	4988	0.5757 -		SFAM01.1	PPB	9/11/2023 14:05	SMPL
O4114-08	MEZYL3	4661	0.5205 -		SFAM01.1	PPB	9/11/2023 14:08	SMPL
O4114-09	MEZYH7	790	-0.1333 -		SFAM01.1	PPB	9/11/2023 14:10	SMPL
O4114-10	MEZYH8	9552	1.3466 -		SFAM01.1	PPB	9/11/2023 14:12	SMPL
O4114-11	MEZYH9	6586	0.8456 -		SFAM01.1	PPB	9/11/2023 14:14	SMPL
O4114-13	MEZYJ0	15120	2.2871 -		SFAM01.1	PPB	9/11/2023 14:17	SMPL
O4114-14	MEZYJ1	394	-0.2002 -		SFAM01.1	PPB	9/11/2023 14:19	SMPL
O4114-15	MEZYJ2	3521	0.328 -		SFAM01.1	PPB	9/11/2023 14:21	SMPL
O4114-16	MEZYG5	681	-0.1517 -		SFAM01.1	PPB	9/11/2023 14:23	SMPL
O4114-17	MEZYG6	193	-0.2342 -		SFAM01.1	PPB	9/11/2023 14:26	SMPL
O4114-18	MEZYG7	25621	4.0607 -		SFAM01.1	PPB	9/11/2023 14:28	SMPL
O4114-19	MEZYGX0	370	-0.2043 -		SFAM01.1	PPB	9/11/2023 14:30	SMPL
O4114-20	MEZYGX3	631	-0.1602 -		SFAM01.1	PPB	9/11/2023 14:32	SMPL
CCV076	CCV076	31007	4.9705 -		SFAM01.1	PPB	9/11/2023 14:35	SMPL
CCB076	CCB076	1511	-0.0115 -		SFAM01.1	PPB	9/11/2023 14:37	SMPL
O4114-21	MEZYGX4	779	-0.1352 -		SFAM01.1	PPB	9/11/2023 14:39	SMPL
CCV077	CCV077	31243	5.0103 -		SFAM01.1	PPB	9/11/2023 14:42	SMPL
CCB077	CCB077	1511	-0.0115 -		SFAM01.1	PPB	9/11/2023 14:44	SMPL