



### Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB127235

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	09/11/2023	15:30
2	CAL2	1	Water	NA	NA	20.5	7.00	09/11/2023	15:31
3	CAL3	1	Water	NA	NA	20.6	10.02	09/11/2023	15:33
4	ICV	1	Water	NA	NA	20.7	7.02	09/11/2023	15:37
5	CCV1	1	Water	NA	NA	20.7	2.01	09/11/2023	15:40
6	O4365-01	1	Water	NA	NA	20.7	9.30	09/11/2023	15:45
7	O4391-01	1	Water	NA	NA	22.3	6.80	09/11/2023	15:50
8	O4391-01DUP	1	Water	NA	NA	22.4	6.81	09/11/2023	15:51
9	CCV2	1	Water	NA	NA	20.7	12.02	09/11/2023	16:00