


LB127239

7471B

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 1.5067e-004

C= 2.9320e-002

Rho= 0.9997970

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.058	0.058	189	0.000	189					14
0.2	0.200	0.227	0.027	1314	0.0 %	1314					-4
2.5	2.500	2.409	-0.091	15793	0.0 %	15793					-1
5.0	5.000	5.030	0.030	33188	0.0 %	33188					-1
7.5	7.500	7.389	-0.111	48848	0.0 %	48848					-1
10.0	10.000	10.087	0.087	66755	0.0 %	66755					-1

LB127239 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 12 Sep 2023 10:36:25

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	189	-	0.0000	7471B	PPB	12 Sep 2023 10:45:45	S	Std
0.2 - 1	50.2	1314	-	0.2000	7471B	PPB	12 Sep 2023 10:48:02	S	Std
2.5 - 1	52.5	15793	-	2.5000	7471B	PPB	12 Sep 2023 10:50:19	S	Std
5.0 - 1	55	33188	-	5.0000	7471B	PPB	12 Sep 2023 10:52:35	S	Std
7.5 - 1	57.5	48848	-	7.5000	7471B	PPB	12 Sep 2023 10:57:26	S	Std
10.0 - 1	510	66755	-	10.0000	7471B	PPB	12 Sep 2023 11:02:15	S	Std
ICV57 - 1	ICV57	27356	4.1510	-	7471B	PPB	12 Sep 2023 11:04:54	U	SMPL
ICB57 - 1	ICB57	58	0.0381	-	7471B	PPB	12 Sep 2023 11:07:09	U	SMPL
CCV78 - 1	CCV78	33102	5.0167	-	7471B	PPB	12 Sep 2023 11:09:24	U	SMPL
CCB78 - 1	CCB78	57	0.0379	-	7471B	PPB	12 Sep 2023 11:11:39	U	SMPL
CRA - 1	CRA	1265	0.2199	-	7471B	PPB	12 Sep 2023 11:13:54	U	SMPL
HighStd - 1	HighStd	67093	10.1381	-	7471B	PPB	12 Sep 2023 11:16:09	U	SMPL
ChkStd - 1	ChkStd	45640	6.9058	-	7471B	PPB	12 Sep 2023 11:18:24	U	SMPL
PB155461BL - 1	PBS	25	0.0331	-	7471B	PPB	12 Sep 2023 11:20:41	U	SMPL
PB155461BS - 1	LCSS	24771	3.7615	-	7471B	PPB	12 Sep 2023 11:25:30	U	SMPL
O4351-02 - 1	CPT101-02	13455	2.0566	-	7471B	PPB	12 Sep 2023 11:27:46	U	SMPL
O4351-04 - 1	CPT101-01	2147	0.3528	-	7471B	PPB	12 Sep 2023 11:30:03	U	SMPL
O4351-06 - 1	B101-01	2076	0.3421	-	7471B	PPB	12 Sep 2023 11:32:20	U	SMPL
O4351-08 - 1	CPT101-10	4183	0.6596	-	7471B	PPB	12 Sep 2023 11:34:36	U	SMPL
O4351-10 - 1	CPT102-01	1354	0.2333	-	7471B	PPB	12 Sep 2023 11:36:51	U	SMPL
CCV79 - 1	CCV79	32481	4.9232	-	7471B	PPB	12 Sep 2023 11:39:06	U	SMPL
CCB79 - 1	CCB79	44	0.0359	-	7471B	PPB	12 Sep 2023 11:41:21	U	SMPL
O4363-01 - 1	72-11940	2531	0.4107	-	7471B	PPB	12 Sep 2023 11:48:32	U	SMPL
O4363-01DUP - 1	72-11940DUP	2398	0.3906	-	7471B	PPB	12 Sep 2023 11:50:59	U	SMPL
O4363-01MS - 1	72-11940MS	27603	4.1882	-	7471B	PPB	12 Sep 2023 11:53:14	U	SMPL
O4363-01MSD - 1	72-11940MSD	26369	4.0023	-	7471B	PPB	12 Sep 2023 11:55:29	U	SMPL
O4363-03 - 1	VNJ-227	7913	1.2216	-	7471B	PPB	12 Sep 2023 11:57:45	U	SMPL
O4363-05 - 1	VNJ-222	5654	0.8812	-	7471B	PPB	12 Sep 2023 12:00:01	U	SMPL
O4363-07 - 1	VNJ-204	3515	0.5589	-	7471B	PPB	12 Sep 2023 12:02:18	U	SMPL
O4367-01 - 1	PILE-1	116347	17.5591	-	7471B	PPB	12 Sep 2023 12:04:34	U	SMPL
O4367-02 - 1	PILE-2	126210	19.0452	-	7471B	PPB	12 Sep 2023 12:06:51	U	SMPL
O4367-03 - 1	PILE-3	250338	37.7474	-	7471B	PPB	12 Sep 2023 12:09:09	U	SMPL
CCV80 - 1	CCV80	32654	4.9492	-	7471B	PPB	12 Sep 2023 12:11:27	U	SMPL
CCB80 - 1	CCB80	6	0.0302	-	7471B	PPB	12 Sep 2023 12:13:45	U	SMPL
O4367-04 - 1	PILE-4	320132	48.2631	-	7471B	PPB	12 Sep 2023 12:16:01	U	SMPL
O4370-01 - 1	EO-02-09082023	1403	0.2407	-	7471B	PPB	12 Sep 2023 12:18:16	U	SMPL
O4371-01 - 1	289	3186	0.5093	-	7471B	PPB	12 Sep 2023 12:20:36	U	SMPL
O4363-01LX5 - 1		472	0.1004	-	7471B	PPB	12 Sep 2023 12:22:51	U	SMPL
O4363-01A - 1		27164	4.1221	-	7471B	PPB	12 Sep 2023 12:25:07	U	SMPL
O4367-01DLX5 - 1		23715	3.6024	-	7471B	PPB	12 Sep 2023 12:27:24	U	SMPL
O4367-02DLX5 - 1		26528	4.0263	-	7471B	PPB	12 Sep 2023 12:29:41	U	SMPL
O4367-03DLX10 - 1		23199	3.5247	-	7471B	PPB	12 Sep 2023 12:31:59	U	SMPL
O4367-04DLX10 - 1		37212	5.6360	-	7471B	PPB	12 Sep 2023 12:34:17	U	SMPL
CCV81 - 1	CCV81	33004	5.0020	-	7471B	PPB	12 Sep 2023 12:36:34	U	SMPL
CCB81 - 1	CCB81	26	0.0332	-	7471B	PPB	12 Sep 2023 12:38:49	U	SMPL