


LB 127267

7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.2046e-004

C= -6.8747e-002

Rho= 0.9997960

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%1)
0.0	0.000	-0.040	-0.040	129	0.000	129	0				17
0.2	0.200	0.187	-0.013	1159	0.0 %	1159					17
2.5	2.500	2.520	0.020	11744	0.0 %	11744					17
5.0	5.000	5.138	0.138	23617	0.0 %	23617					17
7.5	7.500	7.390	-0.110	33833	0.0 %	33833					17
10.0	10.000	10.005	0.005	45697	0.0 %	45697					17

LB127267 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 13 Sep 2023 14:00:16

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	129	-	0.0000	7470A	PPB	13 Sep 2023 14:02:55	S	Std
0.2 - 1	50.2	1159	-	0.2000	7470A	PPB	13 Sep 2023 14:05:11	S	Std
2.5 - 1	52.5	11744	-	2.5000	7470A	PPB	13 Sep 2023 14:07:29	S	Std
5.0 - 1	57.5	23617	-	5.0000	7470A	PPB	13 Sep 2023 14:09:46	S	Std
7.5 - 1	50	33833	-	7.5000	7470A	PPB	13 Sep 2023 14:12:02	S	Std
10.0 - 1		45697	-	10.0000	7470A	PPB	13 Sep 2023 14:14:18	S	Std
ICV61 - 1	ICV61	18430	3.9943	-7470A	PPB	13 Sep 2023 14:19:32	U	SMPL	
ICB61 - 1	ICB61	208	-0.0229	-7470A	PPB	13 Sep 2023 14:21:46	U	SMPL	
CCV95 - 1	CCV95	21705	4.7163	-7470A	PPB	13 Sep 2023 14:24:01	U	SMPL	
CCB95 - 1	CCB95	216	-0.0213	-7470A	PPB	13 Sep 2023 14:26:16	U	SMPL	
CRA - 1	CRA	1062	0.1654	-7470A	PPB	13 Sep 2023 14:28:31	U	SMPL	
HighStd - 1	HighStd	44388	9.7169	-7470A	PPB	13 Sep 2023 14:30:46	U	SMPL	
ChkStd - 1	ChkStd	26514	5.7765	-7470A	PPB	13 Sep 2023 14:33:01	U	SMPL	
PB155528BL - 1	PBW	-52	-0.0802	-7470A	PPB	13 Sep 2023 14:35:17	U	SMPL	
PB155528BS - 1	LCSW	16506	3.5701	-7470A	PPB	13 Sep 2023 14:37:33	U	SMPL	
O4399-02 - 1	72-11930	-46	-0.0789	-7470A	PPB	13 Sep 2023 14:39:49	U	SMPL	
O4399-04 - 1	72-11985	-10	-0.0710	-7470A	PPB	13 Sep 2023 14:42:06	U	SMPL	
O4399-06 - 1	360	73	-0.0527	-7470A	PPB	13 Sep 2023 14:44:22	U	SMPL	
O4400-02 - 1	JOYCE-KILMER-SOIL	31	-0.0619	-7470A	PPB	13 Sep 2023 14:46:39	U	SMPL	
O4401-03 - 1	MH110	20	-0.0643	-7470A	PPB	13 Sep 2023 14:48:54	U	SMPL	
CCV96 - 1	CCV96	22569	4.9067	-7470A	PPB	13 Sep 2023 14:51:09	U	SMPL	
CCB96 - 1	CCB96	228	-0.0185	-7470A	PPB	13 Sep 2023 14:53:24	U	SMPL	
O4401-03DUP - 1	MH110DUP	61	-0.0553	-7470A	PPB	13 Sep 2023 14:55:40	U	SMPL	
O4401-03MS - 1	MH110MS	13404	2.8863	-7470A	PPB	13 Sep 2023 14:57:55	U	SMPL	
O4401-03MSD - 1	MH110MSD	13145	2.8292	-7470A	PPB	13 Sep 2023 15:00:10	U	SMPL	
PB155529BL - 1	PBW	15	-0.0654	-7470A	PPB	13 Sep 2023 15:02:25	U	SMPL	
PB155529BS - 1	LCSW	17006	3.6803	-7470A	PPB	13 Sep 2023 15:04:41	U	SMPL	
O3572-07 - 1	LOD-MDL-WATER-01-QT3-2023	562	0.0551	-7470A	PPB	13 Sep 2023 15:18:47	U	SMPL	
O3572-08 - 1	LOQ-WATER-02-QT3-2023	1365	0.2322	-7470A	PPB	13 Sep 2023 15:23:39	U	SMPL	
O4391-01 - 1	GW-26-WEST-MAIN-STREET	422	0.0243	-7470A	PPB	13 Sep 2023 15:25:54	U	SMPL	
O4391-01DUP - 1	GW-26-WEST-MAIN-STREETDUP	186	-0.0277	-7470A	PPB	13 Sep 2023 15:28:11	U	SMPL	
O4391-01MS - 1	GW-26-WEST-MAIN-STREETMS	14549	3.1387	-7470A	PPB	13 Sep 2023 15:30:26	U	SMPL	
CCV97 - 1	CCV97	23246	5.0560	-7470A	PPB	13 Sep 2023 15:32:41	U	SMPL	
CCB97 - 1	CCB97	182	-0.0286	-7470A	PPB	13 Sep 2023 15:34:57	U	SMPL	
O4391-01MSD - 1	GW-26-WEST-MAIN-STREETMSD	15275	3.2987	-7470A	PPB	13 Sep 2023 15:37:12	U	SMPL	
O4394-11 - 1	FB9/8	-23	-0.0738	-7470A	PPB	13 Sep 2023 15:39:29	U	SMPL	
O4397-09 - 1	FB-9-11	57	-0.0562	-7470A	PPB	13 Sep 2023 15:41:44	U	SMPL	
PB155528TB - 1	PB155528TB	122	-0.0419	-7470A	PPB	13 Sep 2023 15:46:47	U	SMPL	
O4401-03LX5 - 1		-9	-0.0707	-7470A	PPB	13 Sep 2023 15:49:03	U	SMPL	
O4401-03A - 1		17478	3.7844	-7470A	PPB	13 Sep 2023 15:51:19	U	SMPL	
O4391-01LX5 - 1		-7	-0.0703	-7470A	PPB	13 Sep 2023 15:53:36	U	SMPL	
O4391-01A - 1		14375	3.1003	-7470A	PPB	13 Sep 2023 15:55:53	U	SMPL	
CCV98 - 1	CCV98	23037	5.0099	-7470A	PPB	13 Sep 2023 16:01:24	U	SMPL	
CCB98 - 1	CCB98	197	-0.0253	-7470A	PPB	13 Sep 2023 16:03:39	U	SMPL	