



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB127277

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	09/14/2023	07:39
2	CAL2	1	Water	NA	NA	20.7	6.98	09/14/2023	07:40
3	CAL3	1	Water	NA	NA	20.7	10.02	09/14/2023	07:41
4	ICV	1	Water	NA	NA	20.7	7.02	09/14/2023	07:45
5	CCV1	1	Water	NA	NA	20.6	2.01	09/14/2023	07:50
6	O4416-01	1	Water	NA	NA	20.6	6.64	09/14/2023	08:00
7	O4416-01DUP	1	Water	NA	NA	20.7	6.65	09/14/2023	08:01
8	O4416-02	1	Water	NA	NA	20.7	6.64	09/14/2023	08:02
9	CCV2	1	Water	NA	NA	20.7	12.02	09/14/2023	08:05

✓ 127247

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 04416 WorkList ID : 173566 Department : Wet-Chemistry Date : 09-14-2023 07:26:02

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4416-01	A591	Water	pH	Cool 4 deg C	PSEG03	N11	09/13/2023	9040C
O4416-02	A591	Water	pH	Cool 4 deg C	PSEG03	N11	09/13/2023	9040C

Date/Time 09-14-23 07:30
Raw Sample Received by: 70621
Raw Sample Relinquished by: 3065m

Date/Time 09-14-23 10:00
Raw Sample Received by: 3065m
Raw Sample Relinquished by: 70621