



TOTAL VOLATILE SOLIDS SM2540 G

TEMP1 IN: 103 °C 09/14/2023 16:10 TEMP1 OUT: 104 °C 09/15/2023 07:00
TEMP2 IN: 104 °C 09/15/2023 07:30 TEMP2 OUT: 103 °C 09/15/2023 09:00
TEMP3 IN: 540 °C 09/15/2023 09:30 TEMP3 OUT: 540 °C 09/15/2023 11:00
TEMP4 IN: 560 °C 09/15/2023 11:30 TEMP4 OUT: 560 °C 09/15/2023 13:00

Run Number: LB127286

SUPERVISOR: Iwona

ANALYST: jignesh

BalanceID: WC SC-6

OvenID: WC OCEN-1

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (mg/L)
1	LB127286BL	84.7963	84.7963	100	84.7963	84.7963	84.7963	84.7963	0.0000	0
2	O4416-01	75.5454	75.5454	100	75.5780	75.578	75.5656	75.5656	0.0124	124
3	O4416-01DUP	91.3803	91.3803	100	91.4131	91.4131	91.4002	91.4002	0.0129	129

A = Sample Volume (ml)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105°C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105°C (g)

Weight D = C - B

Result % = $\frac{D}{A} \star 1000000$

VB127286

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tvs 04416 WorkList ID : 173570 Department : Wet-Chemistry Date : 09-14-2023 07:30:29

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4416-01	A591	Water	TVS	Cool 4 deg C	PSEG03	N11	09/13/2023	SM2540 G

Date/Time 09.14.27 12.30
Raw Sample Received by: 10/20/21
Raw Sample Relinquished by: CL SM

Date/Time 09.14.23 16.30
Raw Sample Received by: CL SM
Raw Sample Relinquished by: 10/20/21