



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB127292

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	09/14/2023	11:40
2	O4391-01		1	12	100.00	>212.0	>212.0	09/14/2023	12:10
3	O4391-01DUP		1	13	100.00	>212.0	>212.0	09/14/2023	12:40
4	O4416-01		1	14	100.00	>212.0	>212.0	09/14/2023	13:10
5	O4416-02		1	12	100.00	>212.0	>212.0	09/14/2023	13:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66127292

WorkList Name : fp 9-12

WorkList ID : 173492

Department : Wet-Chemistry

Date : 09-12-2023 08:36:42

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4391-01	GW-26-WEST-MAIN-STREET	Water	Flash Point	Cool 4 deg C	PSEG03	N11	09/11/2023	1010B
O4416-01	A591	Water	Flash Point	Cool 4 deg C	PSEG03		09/13/2023	1010B
O4416-02	A591	Water	Flash Point	Cool 4 deg C	PSEG03	N11	09/13/2023	1010B

Date/Time 09/14/2023 14:30
 Raw Sample Received by: RM
 Raw Sample Relinquished by: JH GOC

Date/Time 09/14/2023 14:15
 Raw Sample Received by: JH GOC
 Raw Sample Relinquished by: RM GOC