

LB127302


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

50370



A= 0.0000e+000

B= 1.9808e-004 slope

C= 1.0382e-003 intercept

Rho= 0.9998679

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.035	0.035	171	0.000	171					
0.05	0.050					501	0				
0.20	0.200	0.198	-0.002	995	0.0 %	995					
2.50	2.500	2.495	-0.005	12592	0.0 %	12592					
5.00	5.000	4.897	-0.103	24718	0.0 %	24718					
7.50	7.500	7.596	0.096	38345	0.0 %	38345					
10.0	10.000	9.978	-0.022	50370	0.0 %	50370					

LB127302

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
	0 S0	171	-		0 SFAM01.1	PPB	9/15/2023 9:58	Std
	0.2 S01	995	-		0.2 SFAM01.1	PPB	9/15/2023 10:00	Std
	2.5 S02	12592	-		2.5 SFAM01.1	PPB	9/15/2023 10:03	Std
	5 S03	24718	-		5 SFAM01.1	PPB	9/15/2023 10:05	Std
	7.5 S04	38345	-		7.5 SFAM01.1	PPB	9/15/2023 10:07	Std
	10 S05	50370	-		10 SFAM01.1	PPB	9/15/2023 10:12	Std
ICV064	ICV064	20152	3.9927 -		SFAM01.1	PPB	9/15/2023 10:15	SMPL
ICB064	ICB064	112	0.0232 -		SFAM01.1	PPB	9/15/2023 10:17	SMPL
CCV007	CCV007	24585	4.8708 -		SFAM01.1	PPB	9/15/2023 10:19	SMPL
CCB007	CCB007	86	0.0181 -		SFAM01.1	PPB	9/15/2023 10:21	SMPL
PB155581BL	PBS	63	0.0135 -		SFAM01.1	PPB	9/15/2023 10:24	SMPL
O4176-01	MEZYK2	448	0.0898 -		SFAM01.1	PPB	9/15/2023 10:26	SMPL
O4176-02	MEZYK3	53	0.0115 -		SFAM01.1	PPB	9/15/2023 10:28	SMPL
O4176-03	MEZYK4	1790	0.3556 -		SFAM01.1	PPB	9/15/2023 10:30	SMPL
O4176-05	MEZYC9	202	0.0411 -		SFAM01.1	PPB	9/15/2023 10:33	SMPL
O4176-06	MEZYC9D	91	0.0191 -		SFAM01.1	PPB	9/15/2023 10:35	SMPL
O4176-07	MEZYC9S	13729	2.7205 -		SFAM01.1	PPB	9/15/2023 10:40	SMPL
O4176-08	MEZYD0	109	0.0226 -		SFAM01.1	PPB	9/15/2023 10:42	SMPL
O4176-09	MEZYD1	617	0.1233 -		SFAM01.1	PPB	9/15/2023 10:44	SMPL
O4176-10	MEZYD2	158	0.0323 -		SFAM01.1	PPB	9/15/2023 10:47	SMPL
O4176-11	MEZYD3	2060	0.4091 -		SFAM01.1	PPB	9/15/2023 10:49	SMPL
O4176-12	MEZYD4	7677	1.5217 -		SFAM01.1	PPB	9/15/2023 10:51	SMPL
O4176-13	MEZYW5	3264	0.6476 -		SFAM01.1	PPB	9/15/2023 10:53	SMPL
O4176-15	MEZYD8	423	0.0848 -		SFAM01.1	PPB	9/15/2023 10:56	SMPL
O4176-16	MEZYD9	144	0.0296 -		SFAM01.1	PPB	9/15/2023 10:58	SMPL
O4176-17	MEZYE0	99	0.0206 -		SFAM01.1	PPB	9/15/2023 11:00	SMPL
O4176-18	MEZYJ9	576	0.1151 -		SFAM01.1	PPB	9/15/2023 11:02	SMPL
O4176-19	MEZYK0	240	0.0486 -		SFAM01.1	PPB	9/15/2023 11:05	SMPL
O4176-20	MEZYK1	4313	0.8553 -		SFAM01.1	PPB	9/15/2023 11:07	SMPL
O4176-21	MEZYW8	3688	0.7316 -		SFAM01.1	PPB	9/15/2023 11:09	SMPL
CCV008	CCV008	24695	4.8926 -		SFAM01.1	PPB	9/15/2023 11:12	SMPL
CCB008	CCB008	89	0.0187 -		SFAM01.1	PPB	9/15/2023 11:14	SMPL