



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 9/18/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 13:05
In Date: 09/15/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:30
Out Date: 09/16/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB127307

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O4442-01	MJKTY6	1	1.18	8.67	9.85	7.67	74.9	
O4442-02	MJKTY7	2	1.18	8.44	9.62	7.73	77.6	
O4442-03	MJKTY8	3	1.19	8.66	9.85	7.1	68.2	
O4442-04	MJKTY9	4	1.16	8.53	9.69	6.89	67.2	
O4442-05	MJKTZ0	5	1.15	8.45	9.6	7.25	72.2	
O4442-06	MJKTZ1	6	1.19	8.43	9.62	6.81	66.7	
O4442-07	MJKTZ2	7	1.12	8.77	9.89	6.54	61.8	
O4442-08	MJKTZ3	8	1.17	8.79	9.96	7.23	68.9	
O4442-09	MJKTZ4	9	1.14	8.82	9.96	6.94	65.8	
O4442-10	MJKTZ5	10	1.18	8.79	9.97	7.39	70.6	
O4442-11	MJKTZ6	11	1.16	8.60	9.76	8.53	85.7	
O4442-12	MJKTZ7	12	1.12	8.67	9.79	7.65	75.3	
O4442-13	MJKTZ8	13	1.19	8.72	9.91	7.42	71.4	
O4442-14	MJKTZ9	14	1.11	8.84	9.95	7.58	73.2	
O4442-15	MJKW00	15	1.16	8.62	9.78	7.42	72.6	
O4442-16	MJKW11	16	1.15	8.77	9.92	6.99	66.6	
O4442-17	MJKW12	17	1.13	8.45	9.58	6.61	64.9	
O4442-18	MJKW13	18	1.14	8.82	9.96	7.04	66.9	
O4442-19	MJKW14	19	1.12	8.75	9.87	6.88	65.8	
O4442-20	MJKW14D	20	1.12	8.75	9.87	6.88	65.8	
O4442-21	MJKW14S	21	1.12	8.75	9.87	6.88	65.8	
O4442-22	MJKW15	22	1.19	8.79	9.98	7.51	71.9	
O4460-01	DEL-2-RIGHT-SIDE	23	1.15	8.39	9.54	8.62	89.0	
O4460-03	DEL-2-RIGHT-SIDE-EPH2	24	1.19	8.44	9.63	8.59	87.7	
O4460-04	DEL-1-LEFT-SIDE	25	1.17	8.77	9.94	8.86	87.7	
O4460-06	DEL-1-LEFT-SIDE-EPH2	26	1.19	8.60	9.79	8.73	87.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$