



PERCENT SOLID

Supervisor: Sohil
Analyst: jignesh
Date: 9/21/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 16:35
In Date: 09/15/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:45
Out Date: 09/16/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLIDS-OVEN

QC:LB127314

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O4417-01	EZYF0	1	1.18	8.53	9.71	9.08	92.6	
O4417-02	EZYF1	2	1.15	8.49	9.64	8.95	91.9	
O4417-03	EZYF2	3	1.12	8.65	9.77	9.21	93.5	
O4417-04	EZYF3	4	1.18	8.43	9.61	8.22	83.5	
O4417-05	EZYF4	5	1.18	8.77	9.95	8.92	88.3	
O4417-06	EZYF5	6	1.19	8.77	9.96	8.59	84.4	
O4417-10	VHBLK001	7	1.00	1.00	2.00	2.00	100.0	VHBLK
O4417-11	EZYM9	8	1.19	8.54	9.73	8.91	90.4	
O4417-12	EZYM9MS	9	1.19	8.54	9.73	8.91	90.4	
O4417-13	EZYM9MSD	10	1.19	8.54	9.73	8.91	90.4	
O4417-14	EZYN0	11	1.19	8.45	9.64	8.88	91.0	
O4417-15	EZYN1	12	1.14	8.80	9.94	9.43	94.2	
O4417-16	EZYN2	13	1.18	8.79	9.97	8.81	86.8	
O4417-17	EZYN3	14	1.19	8.72	9.91	8.38	82.5	
O4417-18	EZYN4	15	1.19	8.65	9.84	7.95	78.2	
O4417-19	EZYX1	16	1.16	8.61	9.77	8.31	83.0	
O4417-20	EZYE4	17	1.14	8.83	9.97	9.06	89.7	
O4417-21	EZYE5	18	1.12	8.81	9.93	9.14	91.0	
O4417-22	EZYE6	19	1.13	8.64	9.77	8.88	89.7	
O4417-23	EZYE7	20	1.13	8.64	9.77	8.19	81.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$