



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB127318

Slope : 99.2

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	09/15/2023	15:15
2	CAL2	1	Water	NA	NA	20.5	6.99	09/15/2023	15:16
3	CAL3	1	Water	NA	NA	20.6	10.03	09/15/2023	15:19
4	ICV	1	Water	NA	NA	20.7	7.01	09/15/2023	15:20
5	CCV1	1	Water	NA	NA	20.7	2.01	09/15/2023	15:22
6	O4460-02	1	Solid	20.02	20	20.4	8.99	09/15/2023	15:25
7	O4460-02DUP	1	Solid	20.03	20	20.5	9.00	09/15/2023	15:26
8	O4460-05	1	Solid	20.04	20	20.4	8.40	09/15/2023	15:30
9	CCV2	1	Water	NA	NA	20.6	12.02	09/15/2023	15:35

WORKLIST(Hardcopy Internal Chain)

127318

WorkList Name : CORROSIVITY O4460 WorkList ID : 173645 Department : Wet-Chemistry Date : 09-15-2023 14:52:01

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4460-02	DEL-2-RIGHT-SIDE	Solid	Corrosivity	Cool 4 deg C	PSEG03	N11	09/15/2023	9045D
O4460-05	DEL-1-LEFT-SIDE	Solid	Corrosivity	Cool 4 deg C	PSEG03	N11	09/15/2023	9045D

Date/Time 09-15-23 15:10
 Raw Sample Received by: SQ (COC)
 Raw Sample Relinquished by: JTC (SIN)

Date/Time 09-15-23 17:00
 Raw Sample Received by: JTC (SIN)
 Raw Sample Relinquished by: SQ (COC)