



Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB127325

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	09/18/2023	07:00
2	CAL2	1	Water	NA	NA	20.7	7.01	09/18/2023	07:01
3	CAL3	1	Water	NA	NA	20.6	10.04	09/18/2023	07:05
4	ICV	1	Water	NA	NA	20.6	7.00	09/18/2023	07:06
5	CCV1	1	Water	NA	NA	20.7	2.01	09/18/2023	07:10
6	O4349-21	1	Water	NA	NA	22.6	6.71	09/18/2023	07:15
7	O4454-01	1	Water	NA	NA	21.3	8.98	09/18/2023	07:25
8	O4455-01	1	Water	NA	NA	20.3	5.60	09/18/2023	07:30
9	O4455-02	1	Water	NA	NA	20.1	5.50	09/18/2023	07:33
10	O4455-03	1	Water	NA	NA	20.3	5.57	09/18/2023	07:35
11	O4455-03DUP	1	Water	NA	NA	20.4	5.58	09/18/2023	07:36
12	CCV2	1	Water	NA	NA	20.7	12.02	09/18/2023	07:40

WORKLIST(Hardcopy Internal Chain)

12323

WorkList Name : ph 0455

WorkList ID : 173664

Department : Wet-Chemistry

Date : 09-18-2023 06:17:26

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4349-21	PT-PH-WP	Water	pH	Cool 4 deg C	CHEM02	QA Of	09/05/2023	9040C
O4454-01	33809	Water	pH	Cool 4 deg C	PSEG03	D22	09/15/2023	9040C
O4455-01	CELL-A-POOL-WATER	Water	pH	Cool 4 deg C	STER07	D22	09/12/2023	9040C
O4455-02	CELL-B/C-POOL-WATER	Water	pH	Cool 4 deg C	STER07	D22	09/12/2023	9040C
O4455-03	CELL-D-POOL-WATER	Water	pH	Cool 4 deg C	STER07	D22	09/12/2023	9040C

Date/Time

9-18-23 06:30

Raw Sample Received by:

30 (500)

Raw Sample Relinquished by:

30 (500)

Date/Time

09.18.23

Raw Sample Received by:

30 (500)

Raw Sample Relinquished by:

30 (500)