



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB127326

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	09/18/2023	06:37
2	CAL2	1	Water	NA	NA	20.5	7.00	09/18/2023	06:38
3	CAL3	1	Water	NA	NA	20.5	10.02	09/18/2023	06:39
4	ICV	1	Water	NA	NA	20.6	7.00	09/18/2023	06:40
5	CCV1	1	Water	NA	NA	20.7	2.01	09/18/2023	06:45
6	O4454-02	1	Solid	20.02	20	20.6	6.80	09/18/2023	06:50
7	O4454-02DUP	1	Solid	20.03	20	20.7	6.81	09/18/2023	06:51
8	O4454-03	1	Solid	20.04	20	20.9	6.37	09/18/2023	07:00
9	O4476-01	1	Solid	20.02	20	21.1	5.84	09/18/2023	07:10
10	O4476-02	1	Solid	20.03	20	20.6	5.70	09/18/2023	07:15
11	CCV2	1	Water	NA	NA	20.7	12.02	09/18/2023	07:20

WORKLIST(Hardcopy Internal Chain)

2127326

WorkList Name : ph 04476 WorkList ID : 173665 Department : Wet-Chemistry Date : 09-18-2023 06:18:58

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4454-02	B 34624	Solid	pH	Cool 4 deg C	PSEG03	D22	09/15/2023	9045D
O4454-03	B 34560	Solid	pH	Cool 4 deg C	PSEG03	D22	09/15/2023	9045D
O4476-01	A 60087-60086	Solid	pH	Cool 4 deg C	PSEG03	D31	09/15/2023	9045D
O4476-02	A 60084	Solid	pH	Cool 4 deg C	PSEG03	D31	09/15/2023	9045D

Date/Time 09-18-23 06:30

Raw Sample Received by: 2010001

Raw Sample Relinquished by: 2010001

Date/Time 09-18-23 10:00

Raw Sample Received by: 2010001

Raw Sample Relinquished by: 2010001