



Analytical Summary Report

Analysis Method: SM 4500-H B

Parameter: pH

Run Number: LB127327

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	09/18/2023	07:00
2	CAL2	1	Water	NA	NA	20.7	7.01	09/18/2023	07:01
3	CAL3	1	Water	NA	NA	20.6	10.04	09/18/2023	07:05
4	ICV	1	Water	NA	NA	20.6	7.00	09/18/2023	07:06
5	CCV1	1	Water	NA	NA	20.7	2.01	09/18/2023	07:10
6	O4349-20	1	Water	NA	NA	22.6	6.71	09/18/2023	07:15
7	CCV2	1	Water	NA	NA	20.7	12.02	09/18/2023	07:40

WORKLIST(Hardcopy Internal Chain)

12/23/23

WorkList Name : ph w 04349

WorkList ID : 173666

Department : Wet-Chemistry

Date : 09-18-2023 06:19:55

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4349-20	PT-PH-WP	Water	pH	Cool 4 deg C	CHEM02	QA 01	09/05/2023	SM 4500-H B

Date/Time 09.18.23 06:30
 Raw Sample Received by: JH (QC)
 Raw Sample Relinquished by: SJ (QA office)

Date/Time 09.18.23 10:00
 Raw Sample Received by: N/A
 Raw Sample Relinquished by: JH (QC)