

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	(%)
0.0	0.000	0.006	0.006	112	0.000	112					-
0.2	0.200	0.201	0.001	1211	0.0 %	1211					1
2.5	2.500	2.461	-0.039	13936	0.0 %	13936					-2
5.0	5.000	5.170	0.170	29191	0.0 %	29191					3
7.5	7.500	7.282	-0.218	41082	0.0 %	41082					-3
10.0	10.000	10.080	0.080	56841	0.0 %	56841					1

LB127330 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 18 Sep 2023 10:04:29

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	So	112	-	0.0000	7471B	PPB	18 Sep 2023 10:22:54	S	Std
0.2 - 1	SD-2	1211	-	0.2000	7471B	PPB	18 Sep 2023 10:25:11	S	Std
2.5 - 1	SD-5	13936	-	2.5000	7471B	PPB	18 Sep 2023 10:27:28	S	Std
5.0 - 1	SD-5	29191	-	5.0000	7471B	PPB	18 Sep 2023 10:29:45	S	Std
7.5 - 1	SD-5	41082	-	7.5000	7471B	PPB	18 Sep 2023 10:32:03	S	Std
10.0 - 1	SD-5	56841	-	10.0000	7471B	PPB	18 Sep 2023 10:34:18	S	Std
ICV67 - 1	ICV67	22350	3.9551	-	7471B	PPB	18 Sep 2023 10:37:09	U	SMPL
ICB67 - 1	ICB67	41	-0.0065	-	7471B	PPB	18 Sep 2023 10:39:24	U	SMPL
CCV09 - 1	CCV09	29157	5.1639	-	7471B	PPB	18 Sep 2023 10:41:39	U	SMPL
CCB09 - 1	CCB09	20	-0.0102	-	7471B	PPB	18 Sep 2023 10:43:53	U	SMPL
CRA - 1	CRA	1155	0.1913	-	7471B	PPB	18 Sep 2023 10:46:09	U	SMPL
HighStd - 1	HighStd	56358	9.9943	-	7471B	PPB	18 Sep 2023 10:48:23	U	SMPL
ChkStd - 1	ChkStd	38737	6.8652	-	7471B	PPB	18 Sep 2023 10:50:38	U	SMPL
PB155626BL - 1	PBS	-23	-0.0179	-	7471B	PPB	18 Sep 2023 10:52:55	U	SMPL
O4436-04 - 1	WC-1	4265	0.7436	-	7471B	PPB	18 Sep 2023 10:57:28	U	SMPL
O4436-02 - 1	B103-01	317	0.0425	-	7471B	PPB	18 Sep 2023 10:59:44	U	SMPL
O4436-04 - 1	B103-05	820	0.1318	-	7471B	PPB	18 Sep 2023 11:02:01	U	SMPL
O4436-06 - 1	CPT103-05	696	0.1098	-	7471B	PPB	18 Sep 2023 11:04:18	U	SMPL
O4436-08 - 1	CPT103-02	782	0.1251	-	7471B	PPB	18 Sep 2023 11:06:33	U	SMPL
PB155626BS - 1	LCSS	20348	3.5996	-	7471B	PPB	18 Sep 2023 11:14:18	U	SMPL
CCV10 - 1	CCV10	30865	5.4672	-	7471B	PPB	18 Sep 2023 11:26:34	U	SMPL
CCB10 - 1	CCB10	36	-0.0074	-	7471B	PPB	18 Sep 2023 11:28:49	U	SMPL
O4436-10 - 1	B102-08	14419	2.5468	-	7471B	PPB	18 Sep 2023 11:31:04	U	SMPL
O4436-12 - 1	B102-11	2818	0.4866	-	7471B	PPB	18 Sep 2023 11:33:19	U	SMPL
O4436-14 - 1	CPT102-23	756	0.1205	-	7471B	PPB	18 Sep 2023 11:35:34	U	SMPL
O4436-16 - 1	CPT102-16	781	0.1249	-	7471B	PPB	18 Sep 2023 11:37:50	U	SMPL
O4436-18 - 1	B102-10	3407	0.5912	-	7471B	PPB	18 Sep 2023 11:40:05	U	SMPL
O4436-20 - 1	CPT102-15	420	0.0608	-	7471B	PPB	18 Sep 2023 11:42:22	U	SMPL
O4441-02 - 1	72-11976	11847	2.0900	-	7471B	PPB	18 Sep 2023 11:44:38	U	SMPL
O4445-01 - 1	TR-05-09142023	2937	0.5078	-	7471B	PPB	18 Sep 2023 11:46:54	U	SMPL
O4446-01 - 1	EO-01-091423	6446	1.1309	-	7471B	PPB	18 Sep 2023 11:49:12	U	SMPL
O4446-01DUP - 1	EO-01-091423DUP	5601	0.9808	-	7471B	PPB	18 Sep 2023 11:51:30	U	SMPL
CCV11 - 1	CCV11	29953	5.3053	-	7471B	PPB	18 Sep 2023 11:58:53	U	SMPL
CCB11 - 1	CCB11	64	-0.0024	-	7471B	PPB	18 Sep 2023 12:01:08	U	SMPL
O4446-01MS - 1	EO-01-091423MS	25844	4.5756	-	7471B	PPB	18 Sep 2023 12:05:38	U	SMPL
O4446-01MSD - 1	EO-01-091423MSD	26312	4.6587	-	7471B	PPB	18 Sep 2023 12:07:52	U	SMPL
O4446-01LX5 - 1		1205	0.2002	-	7471B	PPB	18 Sep 2023 12:10:08	U	SMPL
O4446-01A - 1		28732	5.0885	-	7471B	PPB	18 Sep 2023 12:12:23	U	SMPL
CCV12 - 1	CCV12	29376	5.2028	-	7471B	PPB	18 Sep 2023 12:17:22	U	SMPL
CCB12 - 1	CCB12	59	-0.0033	-	7471B	PPB	18 Sep 2023 12:19:38	U	SMPL