


LB127337

7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.0154e-004

C= -2.2295e-002

Rho= 0.9998410

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/01)
0.0	0.000	-0.011	-0.011	54	0.000	54	0				5
0.2	0.200	0.211	0.011	1158	0.0 %	1158					-3
2.5	2.500	2.416	-0.084	12100	0.0 %	12100					2
5.0	5.000	5.111	0.111	25471	0.0 %	25471					1
7.5	7.500	7.541	0.041	37528	0.0 %	37528					-1
10.0	10.000	9.932	-0.068	49389	0.0 %	49389					

LB127337 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 18 Sep 2023 12:32:25

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	54	-	0.0000	7470A	PPB	18 Sep 2023 12:36:33	S	Std
0.2 - 1	50.2	1158	-	0.2000	7470A	PPB	18 Sep 2023 12:38:50	S	Std
2.5 - 1	52.5	12100	-	2.5000	7470A	PPB	18 Sep 2023 12:41:06	S	Std
5.0 - 1	55	25471	-	5.0000	7470A	PPB	18 Sep 2023 12:43:23	S	Std
7.5 - 1	57.5	37528	-	7.5000	7470A	PPB	18 Sep 2023 12:45:39	S	Std
10.0 - 1	51.0	49389	-	10.0000	7470A	PPB	18 Sep 2023 12:47:55	S	Std
ICV68 - 1	ICV68	19513	3.9104	-	7470A	PPB	18 Sep 2023 12:50:52	U	SMPL
ICB68 - 1	ICB68	-66	-0.0356	-	7470A	PPB	18 Sep 2023 12:53:06	U	SMPL
CCV13 - 1	CCV13	25586	5.1343	-	7470A	PPB	18 Sep 2023 12:58:11	U	SMPL
CCB13 - 1	CCB13	-77	-0.0378	-	7470A	PPB	18 Sep 2023 13:00:25	U	SMPL
CRA - 1	CRA	1149	0.2093	-	7470A	PPB	18 Sep 2023 13:02:40	U	SMPL
HighStd - 1	HighStd	48623	9.7773	-	7470A	PPB	18 Sep 2023 13:04:55	U	SMPL
ChkStd - 1	ChkStd	33008	6.6302	-	7470A	PPB	18 Sep 2023 13:07:10	U	SMPL
PB155629BL - 1	PBW	-9	-0.0241	-	7470A	PPB	18 Sep 2023 13:11:31	U	SMPL
PB155629BS - 1	LCSW	18641	3.7346	-	7470A	PPB	18 Sep 2023 13:13:47	U	SMPL
O4409-02 - 1	WC-1(0-2)COMP	578	0.0942	-	7470A	PPB	18 Sep 2023 13:16:03	U	SMPL
O4409-02DUP - 1	WC-1(0-2)COMP DUP	636	0.1059	-	7470A	PPB	18 Sep 2023 13:18:20	U	SMPL
O4409-02MS - 1	WC-1(0-2)COMPMS	17430	3.4906	-	7470A	PPB	18 Sep 2023 13:20:37	U	SMPL
O4409-02MSD - 1	WC-1(0-2)COMPMSD	17418	3.4882	-	7470A	PPB	18 Sep 2023 13:22:53	U	SMPL
O4409-05 - 1	WC-2(0-2)COMP	737	0.1262	-	7470A	PPB	18 Sep 2023 13:25:08	U	SMPL
CCV14 - 1	CCV14	25224	5.0614	-	7470A	PPB	18 Sep 2023 13:27:23	U	SMPL
CCB14 - 1	CCB14	-78	-0.0380	-	7470A	PPB	18 Sep 2023 13:29:39	U	SMPL
O4409-08 - 1	WC-3(0-2)COMP	739	0.1266	-	7470A	PPB	18 Sep 2023 13:31:53	U	SMPL
O4409-11 - 1	WC-4(0-6)COMP	826	0.1442	-	7470A	PPB	18 Sep 2023 13:34:09	U	SMPL
O4409-14 - 1	WC-5(0-6)COMP	781	0.1351	-	7470A	PPB	18 Sep 2023 13:36:24	U	SMPL
O4409-17 - 1	WC-6(0-6)COMP	831	0.1452	-	7470A	PPB	18 Sep 2023 13:38:40	U	SMPL
O4420-01 - 1	091223JE08-TCLP	737	0.1262	-	7470A	PPB	18 Sep 2023 13:40:55	U	SMPL
O4420-02 - 1	091223JE09-TCLP	1041	0.1875	-	7470A	PPB	18 Sep 2023 13:43:11	U	SMPL
O4420-03 - 1	091223JE10-TCLP	10532	2.1003	-	7470A	PPB	18 Sep 2023 13:45:28	U	SMPL
O4420-04 - 1	091223JE11-TCLP	920	0.1631	-	7470A	PPB	18 Sep 2023 13:47:45	U	SMPL
CCV15 - 1	CCV15	25481	5.1132	-	7470A	PPB	18 Sep 2023 13:50:02	U	SMPL
CCB15 - 1	CCB15	-76	-0.0376	-	7470A	PPB	18 Sep 2023 13:52:17	U	SMPL
PB155553TB - 1		9	-0.0205	-	7470A	PPB	18 Sep 2023 13:57:09	U	SMPL
O4409-02LX5 - 1		150	0.0079	-	7470A	PPB	18 Sep 2023 13:59:24	U	SMPL
O4409-02A - 1		18191	3.6439	-	7470A	PPB	18 Sep 2023 14:01:39	U	SMPL
CCV16 - 1	CCV16	25305	5.0777	-	7470A	PPB	18 Sep 2023 14:03:55	U	SMPL
CCB16 - 1	CCB16	-73	-0.0370	-	7470A	PPB	18 Sep 2023 14:06:10	U	SMPL