



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB127338

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.7	09/18/2023	09:40
2	O4454-01		1	14	100.00	>212.0	>212.0	09/18/2023	10:10
3	O4454-01DUP		1	13	100.00	>212.0	>212.0	09/18/2023	10:40
4	O4479-01		1	12	100.00	>212.0	>212.0	09/18/2023	11:10
5	O4479-02		1	12	100.00	>212.0	>212.0	09/18/2023	11:40
6	O4479-03		1	14	100.00	>212.0	>212.0	09/18/2023	12:10
7	O4479-04		1	14	100.00	>212.0	>212.0	09/18/2023	12:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66127338

WorkList Name : fp9-15

WorkList ID : 173616

Department : Wet-Chemistry

Date : 09-15-2023 08:11:15

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4454-01 B	33809	Water	Flash Point	Cool 4 deg C	PSEG03	D22	09/15/2023	1010B
O4479-01 D	1026	Water	Flash Point	Cool 4 deg C	PSEG03	D31	09/15/2023	1010B
O4479-02 D	1027	Water	Flash Point	Cool 4 deg C	PSEG03	D31	09/15/2023	1010B
O4479-03 B	1695	Water	Flash Point	Cool 4 deg C	PSEG03	D31	09/15/2023	1010B
O4479-04 B	1030	Water	Flash Point	Cool 4 deg C	PSEG03	D31	09/15/2023	1010B

Date/Time 09/18/2023 09:30
 Raw Sample Received by: Rm cws
 Raw Sample Relinquished by: JAC

Date/Time 09/18/2023 13:15
 Raw Sample Received by: JAC
 Raw Sample Relinquished by: Rm cws