


LB127352

245.1

INSTRUMENT 1A:CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	I/D
0.0	0.000	-0.008	-0.008	218	0.000	218					
0.05	0.050	0.041	-0.009	451	0.0 %	451	0				-18
0.2	0.200	0.204	0.004	1234	0.0 %	1234					2
2.5	2.500	2.523	0.023	12328	0.0 %	12328					---
5.0	5.000	5.045	0.045	24398	0.0 %	24398					---
7.5	7.500	7.398	-0.102	35654	0.0 %	35654					---
10.0	10.000	10.047	0.047	48331	0.0 %	48331					0

LB127352 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 19 Sep 2023 10:01:12

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	218	-	0.000	245.1	PPB	19 Sep 2023 10:33:40	S	Std
0.05 - 1	50.05	451	-	0.050	245.1	PPB	19 Sep 2023 10:35:56	S	Std
0.2 - 1	51.2	1234	-	0.200	245.1	PPB	19 Sep 2023 10:38:13	S	Std
2.5 - 1	52.5	12328	-	2.500	245.1	PPB	19 Sep 2023 10:40:30	S	Std
5.0 - 1	55	24398	-	5.000	245.1	PPB	19 Sep 2023 10:42:46	S	Std
7.5 - 1	57.5	35854	-	7.500	245.1	PPB	19 Sep 2023 10:45:02	S	Std
10.0 - 1	51.6	48331	-	10.000	245.1	PPB	19 Sep 2023 11:24:11	S	Std
ICV69 - 1	ICV69	19637	4.0503	-245.1		PPB	19 Sep 2023 11:27:11	U	SMPL
ICB69 - 1	ICB69	59	-0.0411	-245.1		PPB	19 Sep 2023 11:29:25	U	SMPL
CCV16 - 1	CCV16	24664	5.1009	-245.1		PPB	19 Sep 2023 11:31:40	U	SMPL
CCB16 - 1	CCB16	53	-0.0424	-245.1		PPB	19 Sep 2023 11:33:55	U	SMPL
CRA - 1	CRA	459	0.0425	-245.1		PPB	19 Sep 2023 11:41:14	U	SMPL
HighStd - 1	HighStd	49054	10.1979	-245.1		PPB	19 Sep 2023 11:43:28	U	SMPL
ChkStd - 1	ChkStd	33630	6.9746	-245.1		PPB	19 Sep 2023 11:45:43	U	SMPL
PB155665BL - 1	PBW	17	-0.0499	-245.1		PPB	19 Sep 2023 11:47:59	U	SMPL
PB155665BS - 1	LCSW	17763	3.6587	-245.1		PPB	19 Sep 2023 11:50:15	U	SMPL
O3572-10 - 1	LOD-MDL-WATER-04-QT3-2023	425	0.0354	-245.1		PPB	19 Sep 2023 11:52:31	U	SMPL
O3572-11 - 1	LOQ-WATER-05-QT3-2023	1062	0.1685	-245.1		PPB	19 Sep 2023 11:54:47	U	SMPL
O4347-08 - 1	PT-HG-WP	109027	22.7312	-245.1		PPB	19 Sep 2023 12:01:38	U	SMPL
CCV17 - 1	CCV17	23529	4.8637	-245.1		PPB	19 Sep 2023 12:03:56	U	SMPL
CCB17 - 1	CCB17	50	-0.0430	-245.1		PPB	19 Sep 2023 12:06:14	U	SMPL
O4347-08DLX5 - 1	PT-HG-WP	24415	5.0488	-245.1		PPB	19 Sep 2023 12:16:41	U	SMPL
CCV18 - 1	CCV18	23400	4.8367	-245.1		PPB	19 Sep 2023 12:21:27	U	SMPL
CVB18 - 1	CVB18	72	-0.0384	-245.1		PPB	19 Sep 2023 12:23:44	U	SMPL