



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB127354

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	09/19/2023	09:40
2	O4490-06		1	12	100.00	>212.0	>212.0	09/19/2023	10:10
3	O4490-06DUP		1	14	100.00	>212.0	>212.0	09/19/2023	10:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB127354

WorkList Name : FP-S-9-18

WorkList ID : 173702

Department : Wet-Chemistry

Date : 09-18-2023 17:08:11

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4490-06 A	103	Solid	Flash Point	Cool 4 deg C	PSEG03	D51	09/18/2023	1010B

Date/Time 09/19/2023 09:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 09/19/2023 11:20
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]