

LB129567


7470A

INSTRUMENT ID: PV1

Linear

 μ Abs.:

51856



A= 0.0000e+000

B= 1.9434e-004

C= -6.7670e-003

Rho= 0.9996715

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/611
0.0	0.000	0.069	0.069	388	0.000	388	0				-8
0.2	0.200	0.184	-0.016	984	0.0 %	984					-4
2.5	2.500	2.408	-0.092	12426	0.0 %	12426					-2
5.0	5.000	5.116	0.116	26358	0.0 %	26358					-2
7.5	7.500	7.352	-0.148	37866	0.0 %	37866					1
10.0	10.000	10.071	0.071	51856	0.0 %	51856					

LB127367 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Sep 2023 15:52:47

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	90	388	-	0.0000	7470A	PPB	19 Sep 2023 15:54:32	S	Std
0.2 - 1	50.2	984	-	0.2000	7470A	PPB	19 Sep 2023 15:56:48	S	Std
2.5 - 1	52.5	12426	-	2.5000	7470A	PPB	19 Sep 2023 15:59:05	S	Std
5.0 - 1	55	26358	-	5.0000	7470A	PPB	19 Sep 2023 16:01:22	S	Std
7.5 - 1	57.5	37866	-	7.5000	7470A	PPB	19 Sep 2023 16:03:38	S	Std
10.0 - 1	510	51856	-	10.0000	7470A	PPB	19 Sep 2023 16:08:35	S	Std
ICV71 - 1	ICV71	20467	3.9708	-	7470A	PPB	19 Sep 2023 16:13:49	U	SMPL
ICB71 - 1	ICB71	265	0.0447	-	7470A	PPB	19 Sep 2023 16:16:04	U	SMPL
CCV23 - 1	CCV23	26315	5.1073	-	7470A	PPB	19 Sep 2023 16:18:19	U	SMPL
CCB23 - 1	CCB23	267	0.0451	-	7470A	PPB	19 Sep 2023 16:20:34	U	SMPL
CRA - 1	CRA	956	0.1790	-	7470A	PPB	19 Sep 2023 16:22:48	U	SMPL
HighStd - 1	HighStd	51356	9.9738	-	7470A	PPB	19 Sep 2023 16:25:03	U	SMPL
ChkStd - 1	ChkStd	35816	6.9537	-	7470A	PPB	19 Sep 2023 16:27:18	U	SMPL
PB155658BL - 1	PBW	72	0.0072	-	7470A	PPB	19 Sep 2023 16:29:35	U	SMPL
PB155658BS - 1	LCSW	18109	3.5126	-	7470A	PPB	19 Sep 2023 16:31:52	U	SMPL
O4447-02 - 1	002	135	0.0195	-	7470A	PPB	19 Sep 2023 16:34:08	U	SMPL
O4460-02 - 1	DEL-2-RIGHT-SIDE	203	0.0327	-	7470A	PPB	19 Sep 2023 16:36:24	U	SMPL
O4460-05 - 1	DEL-1-LEFT-SIDE	218	0.0356	-	7470A	PPB	19 Sep 2023 16:38:40	U	SMPL
O4460-05DUP - 1	DEL-1-LEFT-SIDEDUP	286	0.0488	-	7470A	PPB	19 Sep 2023 16:40:57	U	SMPL
O4460-05MS - 1	DEL-1-LEFT-SIDEMS	19337	3.7512	-	7470A	PPB	19 Sep 2023 16:43:12	U	SMPL
CCV24 - 1	CCV24	26875	5.2161	-	7470A	PPB	19 Sep 2023 16:45:27	U	SMPL
CCB24 - 1	CCB24	265	0.0447	-	7470A	PPB	19 Sep 2023 16:47:42	U	SMPL
O4460-05MSD - 1	DEL-1-LEFT-SIDEMSD	19978	3.8758	-	7470A	PPB	19 Sep 2023 16:49:58	U	SMPL
PB155591TB - 1	PB155591TB	159	0.0241	-	7470A	PPB	19 Sep 2023 16:52:13	U	SMPL
O4460-05LX5 - 1		48	0.0026	-	7470A	PPB	19 Sep 2023 16:54:29	U	SMPL
O4460-05A - 1		19851	3.8511	-	7470A	PPB	19 Sep 2023 16:56:45	U	SMPL
CCV25 - 1	CCV25	27239	5.2869	-	7470A	PPB	19 Sep 2023 16:59:00	U	SMPL
CCB25 - 1	CCB25	165	0.0253	-	7470A	PPB	19 Sep 2023 17:01:16	U	SMPL