



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB127375

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	09/20/2023	07:25
2	CAL2	1	Water	NA	NA	20.7	7.00	09/20/2023	07:26
3	CAL3	1	Water	NA	NA	20.7	10.03	09/20/2023	07:30
4	ICV	1	Water	NA	NA	20.7	7.01	09/20/2023	07:33
5	CCV1	1	Water	NA	NA	20.5	2.02	09/20/2023	08:45
6	O4506-01	1	Water	NA	NA	22.0	7.48	09/20/2023	08:50
7	O4508-01	1	Water	NA	NA	22.0	6.74	09/20/2023	09:00
8	O4516-01	1	Water	NA	NA	21.6	6.23	09/20/2023	09:02
9	O4516-01DUP	1	Water	NA	NA	21.7	6.24	09/20/2023	09:03
10	CCV2	1	Water	NA	NA	20.6	12.02	09/20/2023	09:10

17927375

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY 04416 WorkList ID : 173752 Department : Wet-Chemistry Date : 09-20-2023 07:12:01

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4506-01	D3209	Water	Corrosivity	Cool 4 deg C	PSEG03	N22	09/19/2023	9040C
O4508-01	2924	Water	Corrosivity	Cool 4 deg C	PSEG03	N11	09/19/2023	9040C
O4516-01	371	Water	Corrosivity	Cool 4 deg C	PSEG03	N11	09/19/2023	9040C

Date/Time 09.20.23 07:15
Raw Sample Received by: SQ (WC)
Raw Sample Relinquished by: CBA

Date/Time 09.20.23 11:00
Raw Sample Received by: CBA
Raw Sample Relinquished by: SQ (WC)