



## Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB127376

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W3033         |
| BUFFER PH 7.00 GREEN 1PINT PK6 | W2981         |
| PH 10.01 BUFFER,COLOR CD 475ML | W3043         |
| buffer solution pH 7 yellow    | W2887         |
| Buffer Solution, PH2 (500ml)   | W3005         |
| pH 12.00 Buffer                | W2940         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  | NA          | NA          | 20.6             | 4.01        | 09/20/2023 | 06:40     |
| 2   | CAL2        | 1  | Water  | NA          | NA          | 20.6             | 7.00        | 09/20/2023 | 06:41     |
| 3   | CAL3        | 1  | Water  | NA          | NA          | 20.6             | 10.03       | 09/20/2023 | 06:42     |
| 4   | ICV         | 1  | Water  | NA          | NA          | 20.7             | 7.01        | 09/20/2023 | 06:45     |
| 5   | CCV1        | 1  | Water  | NA          | NA          | 20.7             | 2.01        | 09/20/2023 | 06:46     |
| 6   | O4474-04    | 1  | Water  | NA          | NA          | 20.4             | 6.96        | 09/20/2023 | 06:50     |
| 7   | O4500-01    | 1  | Water  | NA          | NA          | 20.3             | 4.77        | 09/20/2023 | 07:00     |
| 8   | O4511-01    | 1  | Water  | NA          | NA          | 20.6             | 5.13        | 09/20/2023 | 07:05     |
| 9   | O4511-02    | 1  | Water  | NA          | NA          | 20.6             | 4.76        | 09/20/2023 | 07:10     |
| 10  | O4511-02DUP | 1  | Water  | NA          | NA          | 20.7             | 4.77        | 09/20/2023 | 07:11     |
| 11  | CCV2        | 1  | Water  | NA          | NA          | 20.7             | 12.02       | 09/20/2023 | 07:15     |