



Analytical Summary Report

Analysis Method: 120.1

ANALYST: Iwona

Parameter: Conductance

SUPERVISOR REVIEW BY: Sohil

Run Number: LB127382

Meter ID: WC CONDUCTANCE METER-1

Thermometer ID: WC PH METER-1

Reagent/Standard	Lot/Log #
conductivity standard	WP101736

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB127382BL		1	22.1	0.84	0.89	09/20/2023	12:10
2	LB127382BS	1413	1	22.3	1352.00	1425.51	09/20/2023	12:15
3	O4349-08		1	21.5	561.00	601.19	09/20/2023	12:20

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$