


LB127384

SFAM01.1

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>of 10</i>
0.00	0.000	0.060	0.060	254	0.000	254					
0.05	0.050					501	0				
0.20	0.200	0.210	0.010	1138	0.0 %	1138					
2.50	2.500	2.452	-0.048	14293	0.0 %	14293					
5.00	5.000	4.970	-0.030	29077	0.0 %	29077					
7.50	7.500	7.432	-0.068	43527	0.0 %	43527					
10.0	10.000	10.076	0.076	59043	0.0 %	59043					

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	254	-		0 SFAM01.1	PPB	9/20/2023 14:33	Std
	0.2 S01	1138	-		0.2 SFAM01.1	PPB	9/20/2023 14:35	Std
	2.5 S02	14293	-		2.5 SFAM01.1	PPB	9/20/2023 14:38	Std
	5 S03	29077	-		5 SFAM01.1	PPB	9/20/2023 14:40	Std
	7.5 S04	43527	-		7.5 SFAM01.1	PPB	9/20/2023 14:42	Std
	10 S05	59043	-		10 SFAM01.1	PPB	9/20/2023 14:44	Std
ICV072	ICV072	23068	3.9466 -		SFAM01.1	PPB	9/20/2023 14:47	SMPL
ICB072	ICB072	169	0.0452 -		SFAM01.1	PPB	9/20/2023 14:50	SMPL
CCV026	CCV026	28950	4.9487 -		SFAM01.1	PPB	9/20/2023 14:52	SMPL
CCB026	CCB026	158	0.0433 -		SFAM01.1	PPB	9/20/2023 14:54	SMPL
PB155697BL	PBS697	99	0.0333 -		SFAM01.1	PPB	9/20/2023 14:56	SMPL
O4331-02	MEZYG2	12780	2.1938 -		SFAM01.1	PPB	9/20/2023 14:59	SMPL
O4331-03	MEZYG3	51370	8.7685 -		SFAM01.1	PPB	9/20/2023 15:01	SMPL
O4331-04	MEZYG6	2652	0.4682 -		SFAM01.1	PPB	9/20/2023 15:03	SMPL
O4331-05	MEZYG7	3616	0.6325 -		SFAM01.1	PPB	9/20/2023 15:05	SMPL
O4331-06	MEZYG8	3617	0.6327 -		SFAM01.1	PPB	9/20/2023 15:08	SMPL
O4331-08	MEZYF6	220	0.0539 -		SFAM01.1	PPB	9/20/2023 15:10	SMPL
O4331-09	MEZYF7	254	0.0597 -		SFAM01.1	PPB	9/20/2023 15:12	SMPL
O4331-10	MEZYF8	266	0.0617 -		SFAM01.1	PPB	9/20/2023 15:14	SMPL
O4331-11	MEZYF9	720	0.1391 -		SFAM01.1	PPB	9/20/2023 15:17	SMPL
O4331-12	MEZYG0	1634	0.2948 -		SFAM01.1	PPB	9/20/2023 15:19	SMPL
O4331-13	MEZYG1	2488	0.4403 -		SFAM01.1	PPB	9/20/2023 15:21	SMPL
O4331-14	MEZYW7	266	0.0617 -		SFAM01.1	PPB	9/20/2023 15:23	SMPL
O4331-15	MEZYW7D	242	0.0576 -		SFAM01.1	PPB	9/20/2023 15:26	SMPL
O4331-16	MEZYW7S	14224	2.4398 -		SFAM01.1	PPB	9/20/2023 15:30	SMPL
O4331-17	MEZYL4	4876	0.8472 -		SFAM01.1	PPB	9/20/2023 15:33	SMPL
O4331-18	MEZYL5	879	0.1662 -		SFAM01.1	PPB	9/20/2023 15:35	SMPL
O4331-19	MEZYL6	3083	0.5417 -		SFAM01.1	PPB	9/20/2023 15:37	SMPL
PB155698BL	PBS698	58	0.0263 -		SFAM01.1	PPB	9/20/2023 15:39	SMPL
O4415-01	MJKT2	844	0.1602 -		SFAM01.1	PPB	9/20/2023 15:42	SMPL
CCV027	CCV027	28390	4.8533 -		SFAM01.1	PPB	9/20/2023 15:44	SMPL
CCB027	CCB027	153	0.0425 -		SFAM01.1	PPB	9/20/2023 15:46	SMPL
O4415-02	MJKT3	1197	0.2204 -		SFAM01.1	PPB	9/20/2023 15:49	SMPL
O4415-03	MJKT4	1041	0.1938 -		SFAM01.1	PPB	9/20/2023 15:51	SMPL
O4415-04	MJKT5	1026	0.1912 -		SFAM01.1	PPB	9/20/2023 15:53	SMPL
O4415-05	MJKT6	801	0.1529 -		SFAM01.1	PPB	9/20/2023 15:55	SMPL
O4415-06	MJKT7	2738	0.4829 -		SFAM01.1	PPB	9/20/2023 15:58	SMPL
O4415-07	MJKT8	1651	0.2977 -		SFAM01.1	PPB	9/20/2023 16:00	SMPL
O4415-08	MJKT9	1263	0.2316 -		SFAM01.1	PPB	9/20/2023 16:02	SMPL
O4415-09	MJKTW0	1262	0.2314 -		SFAM01.1	PPB	9/20/2023 16:04	SMPL
O4415-10	MJKTW1	836	0.1589 -		SFAM01.1	PPB	9/20/2023 16:07	SMPL
O4415-11	MJKTW2	3536	0.6189 -		SFAM01.1	PPB	9/20/2023 16:09	SMPL
O4415-12	MJKTW3	6146	1.0635 -		SFAM01.1	PPB	9/20/2023 16:11	SMPL
O4415-13	MJKTW4	11189	1.9227 -		SFAM01.1	PPB	9/20/2023 16:13	SMPL
O4415-14	MJKTW5	3549	0.6211 -		SFAM01.1	PPB	9/20/2023 16:16	SMPL
O4415-15	MJKTW6	1648	0.2972 -		SFAM01.1	PPB	9/20/2023 16:18	SMPL

LB127384

INSTRUMENT ID : CV1

O4415-16	MJKTW7	562	0.1122 -	SFAM01.1	PPB	9/20/2023 16:20 SMPL
O4415-17	MJKTW8	572	0.1139 -	SFAM01.1	PPB	9/20/2023 16:22 SMPL
O4415-18	MJKTW9	676	0.1316 -	SFAM01.1	PPB	9/20/2023 16:25 SMPL
O4415-19	MJKTX0	7718	1.3314 -	SFAM01.1	PPB	9/20/2023 16:27 SMPL
O4415-20	MJKTX7	867	0.1641 -	SFAM01.1	PPB	9/20/2023 16:29 SMPL
O4415-21	MJKTX7D	906	0.1708 -	SFAM01.1	PPB	9/20/2023 16:31 SMPL
CCV028	CCV028	29117	4.9772 -	SFAM01.1	PPB	9/20/2023 16:34 SMPL
CCB028	CCB028	154	0.0427 -	SFAM01.1	PPB	9/20/2023 16:36 SMPL
O4415-22	MJKTX7S	17356	2.9734 -	SFAM01.1	PPB	9/20/2023 16:38 SMPL
PB155699BL	PBS699	-16	0.0137 -	SFAM01.1	PPB	9/20/2023 16:41 SMPL
O4355-19	X1247	449394	HIGH -	SFAM01.1	PPB	9/20/2023 16:43
O4355-19D X10	X1247	63195	10.7832 -	SFAM01.1	PPB	9/20/2023 16:59 SMPL
O4355-19DL2X25	X1247	30847	5.2719 -	SFAM01.1	PPB	9/20/2023 17:08 SMPL
CCV029	CCV029	28734	4.9119 -	SFAM01.1	PPB	9/20/2023 17:10 SMPL
CCB029	CCB029	129	0.0384 -	SFAM01.1	PPB	9/20/2023 17:12 SMPL