



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB127385

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	09/20/2023	13:30
2	O4474-04		1	12	100.00	>212.0	>212.0	09/20/2023	14:00
3	O4506-01		1	14	100.00	>212.0	>212.0	09/20/2023	14:30
4	O4508-01		1	14	100.00	>212.0	>212.0	09/20/2023	15:00
5	O4516-01		1	13	100.00	>212.0	>212.0	09/20/2023	15:30
6	O4516-01DUP		1	12	100.00	>212.0	>212.0	09/20/2023	16:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6127385

WorkList Name : fp-9-19

WorkList ID : 173714

Department : Wet-Chemistry

Date : 09-19-2023 08:49:33

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4474-04	H-2313-WATER	Water	Flash Point	Cool 4 deg C	PSEG03	D22	09/15/2023	1010B
O4506-01	D3209	Water	Flash Point	Cool 4 deg C	PSEG03	N22	09/19/2023	1010B
O4508-01	2924	Water	Flash Point	Cool 4 deg C	PSEG03	N11	09/19/2023	1010B
O4516-01	371	Water	Flash Point	Cool 4 deg C	PSEG03	N11	09/19/2023	1010B

Date/Time 09/20/2023 13:20
 Raw Sample Received by: RM CWL
 Raw Sample Relinquished by: JH CWL

Date/Time 09/20/2023 15:35
 Raw Sample Received by: JH CWL
 Raw Sample Relinquished by: RM CWL