


LB127391

7471B

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 1.4882e-004

C= 5.0760e-004

Rho= 0.9999250

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	g.d
0.0	0.000	0.024	0.024	158	0.000	158					1
0.2	0.200	0.201	0.001	1347	0.0 %	1347					-2
2.5	2.500	2.448	-0.052	16444	0.0 %	16444					1
5.0	5.000	5.063	0.063	34020	0.0 %	34020					-1
7.5	7.500	7.436	-0.064	49963	0.0 %	49963					0
10.0	10.000	10.028	0.028	67385	0.0 %	67385					

LB127391 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 21 Sep 2023 08:11:29

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	158	-	0.0000	7471B	PPB	21 Sep 2023 08:18:53	S	Std
0.2 - 1	50.2	1347	-	0.2000	7471B	PPB	21 Sep 2023 08:21:10	S	Std
2.5 - 1	52.5	16444	-	2.5000	7471B	PPB	21 Sep 2023 08:23:27	S	Std
5.0 - 1	55	34020	-	5.0000	7471B	PPB	21 Sep 2023 08:25:43	S	Std
7.5 - 1	57.5	49963	-	7.5000	7471B	PPB	21 Sep 2023 08:28:00	S	Std
10.0 - 1	510	67385	-	10.0000	7471B	PPB	21 Sep 2023 08:30:17	S	Std
ICV73 - 1	ICV73	26726	3.9777	-	7471B	PPB	21 Sep 2023 08:33:32	U	SMPL
ICB73 - 1	ICB73	93	0.0143	-	7471B	PPB	21 Sep 2023 08:35:46	U	SMPL
CCV30 - 1	CCV30	34097	5.0747	-	7471B	PPB	21 Sep 2023 08:38:02	U	SMPL
CCB30 - 1	CCB30	88	0.0136	-	7471B	PPB	21 Sep 2023 08:40:17	U	SMPL
CRA - 1	CRA	1321	0.1971	-	7471B	PPB	21 Sep 2023 08:42:33	U	SMPL
HighStd - 1	HighStd	67621	10.0635	-	7471B	PPB	21 Sep 2023 08:44:49	U	SMPL
ChkStd - 1	ChkStd	47211	7.0262	-	7471B	PPB	21 Sep 2023 08:47:03	U	SMPL
PB155713BL - 1	PBS	-51	-0.0071	-	7471B	PPB	21 Sep 2023 08:49:21	U	SMPL
PB155713BS - 1	LCSS	25997	3.8693	-	7471B	PPB	21 Sep 2023 08:51:39	U	SMPL
O4474-01 - 1	H-2313-SOIL	14740	2.1940	-	7471B	PPB	21 Sep 2023 08:53:55	U	SMPL
O4499-01 - 1	CHRT-20001	10421	1.5513	-	7471B	PPB	21 Sep 2023 08:56:12	U	SMPL
O4501-01 - 1	TP-1	5688	0.8470	-	7471B	PPB	21 Sep 2023 08:58:30	U	SMPL
O4501-04 - 1	TP-3	16171	2.4070	-	7471B	PPB	21 Sep 2023 09:00:47	U	SMPL
O4501-07 - 1	TP-2	3976	0.5922	-	7471B	PPB	21 Sep 2023 09:03:02	U	SMPL
CCV31 - 1	CCV31	34223	5.0934	-	7471B	PPB	21 Sep 2023 09:05:17	U	SMPL
CCB31 - 1	CCB31	155	0.0236	-	7471B	PPB	21 Sep 2023 09:07:32	U	SMPL
O4512-02 - 1	72-11986	3991	0.5944	-	7471B	PPB	21 Sep 2023 09:09:49	U	SMPL
O4515-01 - 1	WC-1	4797	0.7144	-	7471B	PPB	21 Sep 2023 09:12:04	U	SMPL
O4515-01DUP - 1	WC-1DUP	4616	0.6874	-	7471B	PPB	21 Sep 2023 09:14:19	U	SMPL
O4515-01MS - 1	WC-1MS	32901	4.8967	-	7471B	PPB	21 Sep 2023 09:16:35	U	SMPL
O4515-01MSD - 1	WC-1MSD	30402	4.5248	-	7471B	PPB	21 Sep 2023 09:18:50	U	SMPL
O4515-01LX5 - 1		858	0.1282	-	7471B	PPB	21 Sep 2023 09:21:07	U	SMPL
O4515-01A - 1		32026	4.7665	-	7471B	PPB	21 Sep 2023 09:23:23	U	SMPL
CCV32 - 1	CCV32	34486	5.1326	-	7471B	PPB	21 Sep 2023 09:25:40	U	SMPL
CCB32 - 1	CCB32	-44	-0.0060	-	7471B	PPB	21 Sep 2023 09:27:57	U	SMPL