


LB127893

7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 1.9584e-004

C= 1.7575e-002

Rho= 0.9998536

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Q/P/D
0.0	0.000	0.038	0.038	104	0.000	104					1
0.2	0.200	0.199	-0.001	924	0.0 %	924					1
2.5	2.500	2.456	-0.044	12450	0.0 %	12450					2
5.0	5.000	4.938	-0.062	25126	0.0 %	25126					1
7.5	7.500	7.619	0.119	38815	0.0 %	38815					2
10.0	10.000	9.950	-0.050	50716	0.0 %	50716					1

LB127393 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 21 Sep 2023 10:05:30

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	104	-	0.0000	7470A	PPB	21 Sep 2023 10:06:25	S	Std
0.2 - 1	50.2	924	-	0.2000	7470A	PPB	21 Sep 2023 10:08:42	S	Std
2.5 - 1	52.5	12450	-	2.5000	7470A	PPB	21 Sep 2023 10:10:59	S	Std
5.0 - 1	55	25126	-	5.0000	7470A	PPB	21 Sep 2023 10:13:15	S	Std
7.5 - 1	55.5	38815	-	7.5000	7470A	PPB	21 Sep 2023 10:15:32	S	Std
10.0 - 1	80	50716	-	10.0000	7470A	PPB	21 Sep 2023 10:17:47	S	Std
ICV74 - 1	ICV74	18620	3.6642	-	7470A	PPB	21 Sep 2023 10:20:41	U	SMPL
ICB74 - 1	ICB74	-1	0.0174	-	7470A	PPB	21 Sep 2023 10:22:55	U	SMPL
CCV33 - 1	CCV33	26127	5.1344	-	7470A	PPB	21 Sep 2023 10:25:10	U	SMPL
CCB33 - 1	CCB33	8	0.0191	-	7470A	PPB	21 Sep 2023 10:27:23	U	SMPL
CRA - 1	CRA	857	0.1854	-	7470A	PPB	21 Sep 2023 10:29:40	U	SMPL
HighStd - 1	HighStd	50846	9.9755	-	7470A	PPB	21 Sep 2023 10:31:55	U	SMPL
ChkStd - 1	ChkStd	35955	7.0600	-	7470A	PPB	21 Sep 2023 10:34:11	U	SMPL
PB155683BL - 1	PBW	-34	0.0109	-	7470A	PPB	21 Sep 2023 10:36:27	U	SMPL
PB155683BS - 1	LCSW	19874	3.9098	-	7470A	PPB	21 Sep 2023 10:38:44	U	SMPL
O4474-03 - 1	H-2313-SOIL	-11	0.0154	-	7470A	PPB	21 Sep 2023 10:41:00	U	SMPL
O4485-01 - 1	BUILDING-39-ROOM-C-14-C-13	32446	6.3719	-	7470A	PPB	21 Sep 2023 10:48:16	U	SMPL
O4485-01DUP - 1	BUILDING-39-ROOM-C-14-C-13DUP	34641	6.8018	-	7470A	PPB	21 Sep 2023 10:50:32	U	SMPL
O4485-01MS - 1	BUILDING-39-ROOM-C-14-C-13MS	57282	11.2359	-	7470A	PPB	21 Sep 2023 10:52:49	U	SMPL
O4485-01MSD - 1	BUILDING-39-ROOM-C-14-C-13MSD	54331	10.6580	-	7470A	PPB	21 Sep 2023 10:55:04	U	SMPL
CCV34 - 1	CCV34	26455	5.1986	-	7470A	PPB	21 Sep 2023 10:57:21	U	SMPL
CCB34 - 1	CCB34	-39	0.0099	-	7470A	PPB	21 Sep 2023 10:59:38	U	SMPL
O4490-01 - 1	COMP-1	2	0.0180	-	7470A	PPB	21 Sep 2023 11:01:53	U	SMPL
O4490-02 - 1	COMP-2	2	0.0180	-	7470A	PPB	21 Sep 2023 11:04:08	U	SMPL
O4490-03 - 1	COMP-3	15	0.0205	-	7470A	PPB	21 Sep 2023 11:06:23	U	SMPL
O4490-04 - 1	COMP-4	25	0.0225	-	7470A	PPB	21 Sep 2023 11:08:38	U	SMPL
O4490-05 - 1	B5	155	0.0479	-	7470A	PPB	21 Sep 2023 11:10:54	U	SMPL
O4490-06 - 1	103	302	0.0767	-	7470A	PPB	21 Sep 2023 11:13:10	U	SMPL
O4501-02 - 1	TP-1	183	0.0534	-	7470A	PPB	21 Sep 2023 11:15:27	U	SMPL
O4501-05 - 1	TP-3	230	0.0626	-	7470A	PPB	21 Sep 2023 11:17:44	U	SMPL
O4501-08 - 1	TP-2	296	0.0755	-	7470A	PPB	21 Sep 2023 11:20:03	U	SMPL
O4504-01 - 1	BLDG-39-SECOND-FLOOR-091823JE09-TCLP-C-14-15-A	304	0.0771	-	7470A	PPB	21 Sep 2023 11:22:19	U	SMPL
CCV35 - 1	CCV35	26474	5.2024	-	7470A	PPB	21 Sep 2023 11:24:34	U	SMPL
CCB35 - 1	CCB35	28	0.0231	-	7470A	PPB	21 Sep 2023 11:26:50	U	SMPL
O4504-03 - 1	BLDG-93-FIRST-FLOOR-091823JE10-TCLP-C-71-72-B	11325	2.2355	-	7470A	PPB	21 Sep 2023 11:29:05	U	SMPL
O4506-02 - 1	D3216	-12	0.0152	-	7470A	PPB	21 Sep 2023 11:31:20	U	SMPL
O4512-01 - 1	214	32	0.0238	-	7470A	PPB	21 Sep 2023 11:33:35	U	SMPL
PB155684BL - 1	PBW	26	0.0227	-	7470A	PPB	21 Sep 2023 11:35:50	U	SMPL
PB155684BS - 1	LCSW	18988	3.7363	-	7470A	PPB	21 Sep 2023 11:38:06	U	SMPL
O4506-01 - 1	D3209	117	0.0405	-	7470A	PPB	21 Sep 2023 11:40:22	U	SMPL
O4508-01 - 1	2924	-48	0.0082	-	7470A	PPB	21 Sep 2023 11:42:38	U	SMPL
O4516-01 - 1	371	69	0.0311	-	7470A	PPB	21 Sep 2023 11:44:55	U	SMPL
O4516-01DUP - 1	371DUP	99	0.0370	-	7470A	PPB	21 Sep 2023 11:47:11	U	SMPL
O4516-01MS - 1	371MS	22361	4.3968	-	7470A	PPB	21 Sep 2023 11:49:27	U	SMPL
CCV36 - 1	CCV36	25044	5.1182	-	7470A	PPB	21 Sep 2023 11:51:42	U	SMPL
CCB36 - 1	CCB36	24	0.0223	-	7470A	PPB	21 Sep 2023 11:53:57	U	SMPL
O4516-01MSD - 1	371MSD	21916	4.3097	-	7470A	PPB	21 Sep 2023 11:56:13	U	SMPL
PB155685BL - 1	PBW	-38	0.0101	-	7470A	PPB	21 Sep 2023 11:58:28	U	SMPL
PB155685BS - 1	LCSW	21483	4.2249	-	7470A	PPB	21 Sep 2023 12:00:43	U	SMPL
O4453-11 - 1	FB-9-14	-58	0.0062	-	7470A	PPB	21 Sep 2023 12:02:59	U	SMPL
O4474-04 - 1	H-2313-WATER	1698	0.3501	-	7470A	PPB	21 Sep 2023 12:05:14	U	SMPL
O4474-04DUP - 1	H-2313-WATERDUP	1755	0.3613	-	7470A	PPB	21 Sep 2023 12:07:31	U	SMPL
O4474-04MS - 1	H-2313-WATERMS	24160	4.7492	-	7470A	PPB	21 Sep 2023 12:09:47	U	SMPL
O4474-04MSD - 1	H-2313-WATERMSD	24920	4.8992	-	7470A	PPB	21 Sep 2023 12:12:04	U	SMPL
O4479-03 - 1	1695	-42	0.0093	-	7470A	PPB	21 Sep 2023 12:14:22	U	SMPL
O4479-04 - 1	1030	451	0.1059	-	7470A	PPB	21 Sep 2023 12:16:39	U	SMPL
CCV37 - 1	CCV37	25914	5.0927	-	7470A	PPB	21 Sep 2023 12:18:54	U	SMPL
CCB37 - 1	CCB37	52	0.0276	-	7470A	PPB	21 Sep 2023 12:21:09	U	SMPL
O4491-09 - 1	FB-9-15	47	0.0268	-	7470A	PPB	21 Sep 2023 12:23:25	U	SMPL
PB155659TB - 1		66	0.0305	-	7470A	PPB	21 Sep 2023 12:25:40	U	SMPL
PB155678TB - 1		45	0.0264	-	7470A	PPB	21 Sep 2023 12:27:55	U	SMPL
O4485-01LX5 - 1		7128	1.4136	-	7470A	PPB	21 Sep 2023 12:30:11	U	SMPL
O4485-01A - 1		54434	10.6782	-	7470A	PPB	21 Sep 2023 12:32:27	U	SMPL
O4516-01LX5 - 1		-114	-0.0048	-	7470A	PPB	21 Sep 2023 12:34:43	U	SMPL
O4516-01A - 1		18746	3.6893	-	7470A	PPB	21 Sep 2023 12:37:01	U	SMPL
O4474-04LX5 - 1		299	0.0761	-	7470A	PPB	21 Sep 2023 12:39:16	U	SMPL
O4474-04A - 1		18880	3.7151	-	7470A	PPB	21 Sep 2023 12:41:34	U	SMPL
CCV38 - 1	CCV38	25481	5.0079	-	7470A	PPB	21 Sep 2023 12:43:50	U	SMPL
CCB38 - 1	CCB38	26	0.0227	-	7470A	PPB	21 Sep 2023 12:46:06	U	SMPL
O4485-01DLX2 - 1		18774	3.6944	-	7470A	PPB	21 Sep 2023 12:48:22	U	SMPL
O4485-01DUPDLX2 - 1		17014	3.3497	-	7470A	PPB	21 Sep 2023 12:50:38	U	SMPL
O4485-01MSDLX2 - 1		28867	5.6710	-	7470A	PPB	21 Sep 2023 12:52:54	U	SMPL
O4485-01MSDDLX2 - 1		30030	5.8988	-	7470A	PPB	21 Sep 2023 12:55:10	U	SMPL
O4485-01LDLX10 - 1		3844	0.7704	-	7470A	PPB	21 Sep 2023 12:57:26	U	SMPL
O4485-01ADLX2 - 1		29068	5.7104	-	7470A	PPB	21 Sep 2023 12:59:42	U	SMPL
CCV39 - 1	CCV39	25633	5.0377	-	7470A	PPB	21 Sep 2023 13:01:58	U	SMPL
CCB39 - 1	CCB39	12	0.0199	-	7470A	PPB	21 Sep 2023 13:04:14	U	SMPL