



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB127394

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	09/21/2023	09:15
2	CAL2	1	Water	NA	NA	20.6	7.00	09/21/2023	09:16
3	CAL3	1	Water	NA	NA	20.5	10.02	09/21/2023	09:18
4	ICV	1	Water	NA	NA	20.6	7.01	09/21/2023	09:20
5	CCV1	1	Water	NA	NA	20.6	2.01	09/21/2023	09:25
6	O4537-02	1	Water	NA	NA	20.2	7.03	09/21/2023	09:30
7	O4537-02DUP	1	Water	NA	NA	20.3	7.04	09/21/2023	09:31
8	O4537-04	1	Water	NA	NA	20.6	7.40	09/21/2023	09:35
9	CCV2	1	Water	NA	NA	20.5	12.02	09/21/2023	09:36

WORKLIST(Hardcopy Internal Chain)

127394

WorkList Name : CORROSIONITY O4537 V WorkList ID : 173796 Department : Wet-Chemistry Date : 09-21-2023 08:59:19

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4537-02	COMP-2	Water	Corrosivity	Cool 4 deg C	PSEG03	D21	09/20/2023	9040C
O4537-04	5866	Water	Corrosivity	Cool 4 deg C	PSEG03	D21	09/20/2023	9040C

Date/Time 09-21-23 09:10
 Raw Sample Received by: LP/WCI
 Raw Sample Relinquished by: CP SM

Date/Time 09-21-23 12:00
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: LP/WCI