



Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB127398

Analyst By : jignesh
Supervisor Review By : Iwona
Slope : 99.1
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	09/21/2023	09:15
2	CAL2	1	Water	NA	NA	20.6	7.00	09/21/2023	09:16
3	CAL3	1	Water	NA	NA	20.5	10.02	09/21/2023	09:18
4	ICV	1	Water	NA	NA	20.6	7.01	09/21/2023	09:20
5	CCV1	1	Water	NA	NA	20.6	2.01	09/21/2023	10:20
6	O4522-01	1	Water	NA	NA	20.3	7.37	09/21/2023	10:25
7	O4524-01	1	Water	NA	NA	20.2	5.41	09/21/2023	10:30
8	O4525-01	1	Water	NA	NA	20.3	5.14	09/21/2023	10:33
9	O4540-01	1	Water	NA	NA	21.1	9.08	09/21/2023	10:40
10	O4540-02	1	Water	NA	NA	21.0	9.23	09/21/2023	10:45
11	O4540-03	1	Water	NA	NA	21.0	8.53	09/21/2023	10:50
12	O4540-03DUP	1	Water	NA	NA	21.1	8.54	09/21/2023	10:51
13	CCV2	1	Water	NA	NA	20.7	12.02	09/21/2023	11:00

VB 127398

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O4524 WorkList ID : 173792 Department : Wet-Chemistry Date : 09-21-2023 08:55:30

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4522-01	A 33927	Water	pH	Cool 4 deg C	PSEG03	D11	09/20/2023	9040C
O4524-01	A A-B-CELL-POOL-WATER-3RD-	Water	pH	Cool 4 deg C	STER12	D22	09/19/2023	9040C
O4525-01	A CELL-A	Water	pH	Cool 4 deg C	STER10	D22	09/18/2023	9040C
O4540-01	F CENTRAL-M-R-HEATER	Water	pH	Cool 4 deg C	PSEG03	N11	09/20/2023	9040C
O4540-02	K EMBRIDGE-M-R-HEATER	Water	pH	Cool 4 deg C	PSEG03	N11	09/20/2023	9040C
O4540-03	L TRANSCO-M-R-HEATER	Water	pH	Cool 4 deg C	PSEG03	N11	09/20/2023	9040C

Date/Time 09-21-23 09:10
 Raw Sample Received by: JN (WCI)
 Raw Sample Relinquished by: CP SM

Date/Time 09-21-23 12:00
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JN (WCI)