



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 9/22/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 17:15
In Date: 09/21/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:22
Out Date: 09/22/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB127402

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O4521-01	350	1	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
O4531-01	TP-BCD	2	1.15	8.80	9.95	9.04	89.7	
O4531-03	TP-BCD-EPH	3	1.15	8.80	9.95	9.04	89.7	
O4531-04	TP-A	4	1.16	8.55	9.71	8.59	86.9	
O4531-06	TP-A-EPH	5	1.16	8.55	9.71	8.59	86.9	
O4535-01	WASTE	11	1.17	8.75	9.92	8.67	85.7	
O4536-01	T-1	12	1.19	8.38	9.57	8.78	90.6	
O4536-02	T-2	13	1.11	8.66	9.77	7.95	79.0	
O4536-03	T-3	14	1.15	8.41	9.56	8.74	90.2	
O4536-04	T-4	15	1.12	8.67	9.79	8.93	90.1	
O4536-05	T-5	16	1.15	8.82	9.97	9.17	90.9	
O4537-01	COMP-1	6	1.00	1.00	2.00	2.00	100.0	OIL SAMPLE
O4537-03	COMP-3	7	1.14	8.52	9.66	8.1	81.7	
O4537-05	5873	8	1.00	1.00	2.00	2.00	100.0	debris
O4537-06	5811	9	1.13	8.39	9.52	6.78	67.3	
O4537-07	5841	10	1.00	1.00	2.00	2.00	100.0	debris
O4544-01	091923-01	17	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O4544-02	091923-02	18	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O4544-03	091923-03	19	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O4544-04	091923-04	20	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O4544-05	091923-05	21	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O4544-06	091923-06	22	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O4544-07	091923-07	23	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O4544-08	091923-08-FB	24	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O4544-09	091923-09-FB	25	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O4544-10	091923-10-FB	26	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O4556-01	SS052P-SD01-00-091923	27	1.15	8.80	9.95	7.56	72.8	
O4556-02	SS052P-SD02-00-091923	28	1.15	8.80	9.95	8.4	82.4	



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 9/22/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 17:15
In Date: 09/21/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:22
Out Date: 09/22/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB127402

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O4556-03	SS052P-SD03-00-091923	29	1.18	8.79	9.97	7.6	73.0	
O4556-04	SS051P-SD04-00-091923	30	1.16	8.81	9.97	8.65	85.0	
O4556-05	AT054P-SD03-00-091924	31	1.19	8.47	9.66	7.62	75.9	
O4556-06	FT013P-SD03-00-091924	32	1.13	8.84	9.97	8.38	82.0	
O4556-07	FT013P-SD01-00-091924	33	1.12	8.85	9.97	8.54	83.8	
O4556-08	FT013P-SD02-00-091924	34	1.1	8.71	9.81	7.24	70.5	
O4556-09	AT054P-SD02-091923	35	1.18	8.72	9.9	7.39	71.2	
O4556-10	SS056P-SD01-00-091923	36	1.15	8.41	9.56	6.56	64.3	
O4556-11	SS055P-SD01-00-091923	37	1.14	8.78	9.92	5.74	52.4	
O4556-12	SS055P-SD02-00-091923	38	1.13	8.79	9.92	5.06	44.7	
O4556-13	AT029P-SD01-00-091923	39	1.17	8.54	9.71	6.68	64.5	
O4556-14	SS057P-SD01-00-091923	40	1.12	8.84	9.96	6.17	57.1	
O4556-15	WP021P-SD01-00-091923	41	1.18	8.78	9.96	5.78	52.4	
O4556-16	SS030P-SD03-00-091923	42	1.19	8.79	9.98	8.4	82.0	
O4556-17	SS057P-SD03-00-091923	43	1.15	8.81	9.96	8.64	85.0	
O4556-18	SS057P-SD02-00-091923	44	1.14	8.82	9.96	5.84	53.3	
O4556-19	SS030P-SD02-00-091923	45	1.14	8.49	9.63	6.00	57.2	
O4556-20	DA053P-SD03-00-091923	46	1.16	8.82	9.98	8.5	83.2	
O4556-21	DA053P-SD02-00-091923	47	1.13	8.83	9.96	8.42	82.6	
O4556-22	DA053P-SD01-00-091923	48	1.16	8.83	9.99	7.01	66.3	
O4556-23	SS050P-SD03-00-091923	49	1.1	8.73	9.83	6.5	61.9	
O4556-24	SS050P-SD02-00-091923	50	1.13	8.86	9.99	7.47	71.6	
O4556-25	SS050P-SD01-00-091923	51	1.17	8.60	9.77	7.95	78.8	
O4556-26	SS030P-SD01-00-091923	52	1.19	8.73	9.92	7.52	72.5	
O4556-27	SS009P-SD03-00-091923	53	1.13	8.85	9.98	8.73	85.9	
O4556-28	SS009P-SD02-00-091923	54	1.18	8.73	9.91	8.54	84.3	
O4556-29	SS009P-SD01-00-091923	55	1.12	8.60	9.72	7.91	79.0	
O4556-30	SS051P-SD01-00-091923	56	1.15	8.82	9.97	8.17	79.6	



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 9/22/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 17:15
In Date: 09/21/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:22
Out Date: 09/22/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB127402

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O4556-31	SS051P-SD03-00-091923	57	1.14	8.54	9.68	7.96	79.9	
O4556-32	SS051P-SD02-00-091923	58	1.15	8.80	9.95	8.00	77.8	
O4557-01	FT011P-SD01-00-092023	59	1.16	8.78	9.94	6.34	59.0	
O4558-01	AT024P-SD04-00-091823-TOC	60	1.19	8.78	9.97	8.7	85.5	
O4558-02	AT024P-SD01-00-091823	61	1.19	8.79	9.98	8.35	81.5	
O4558-03	AT024P-SD02-00-091823	62	1.18	8.61	9.79	7.54	73.9	
O4558-04	AT024P-SD03-00-091823	63	1.19	8.77	9.96	6.88	64.9	
O4559-01	AT014P-SD03-00-092023-TOC	64	1.16	8.49	9.65	8.08	81.5	
O4559-03	SS092P-SD04-00-092023	65	1.18	8.79	9.97	9.2	91.2	
O4559-04	AT014P-SD02-00-092023	66	1.13	8.57	9.7	8.27	83.3	
O4559-05	O4559-04MS	67	1.13	8.57	9.7	8.27	83.3	
O4559-06	O4559-04MSD	68	1.13	8.57	9.7	8.27	83.3	
O4560-01	WP021P-SD03-00-091823	69	1.18	8.58	9.76	5.08	45.5	
O4560-02	WP021P-SD02-00-091823	70	1.19	8.80	9.99	8.35	81.4	
O4560-03	FT011P-SD02-00-091823	71	1.1	8.89	9.99	5.79	52.8	
O4560-04	AT029P-SD03-00-091823	72	1.19	8.79	9.98	8.68	85.2	
O4560-05	FT011P-SD03-00-091923	73	1.19	8.52	9.71	6.83	66.2	
O4560-06	SS055P-SD03-00-091823	74	1.16	8.53	9.69	6.28	60.0	
O4560-07	SS055P-SD04-00-091823	75	1.18	8.79	9.97	8.42	82.4	
O4560-08	AT029P-SD04-00-091823	76	1.15	8.83	9.98	6.86	64.7	
O4560-09	FT011P-SD04-00-091823	77	1.19	8.67	9.86	7.67	74.7	
O4563-04	BEL-23-0007	78	1.00	1.00	2.00	2.00	100.0	debris
O4567-01	SE-SOIL-1	79	1.14	8.61	9.75	8.44	84.8	
O4567-02	SE-SOIL-1	80	1.13	8.67	9.8	8.59	86.0	
O4567-03	SE-SOIL-2	81	1.15	8.43	9.58	8.19	83.5	
O4567-04	SE-SOIL-2	82	1.13	8.71	9.84	8.39	83.4	
O4567-05	SE-SOIL-3	83	1.19	8.62	9.81	8.38	83.4	



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 9/22/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 17:15
In Date: 09/21/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:22
Out Date: 09/22/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB127402

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O4567-06	SE-SOIL-3	84	1.15	8.72	9.87	8.47	83.9	
O4567-07	SE-SOIL-4	85	1.19	8.54	9.73	9.44	96.6	
O4567-08	SE-SOIL-4	86	1.13	8.70	9.83	9.12	91.8	
O4567-09	SE-SOIL-5	87	1.15	8.44	9.59	8.15	82.9	
O4567-10	SE-SOIL-5	88	1.18	8.79	9.97	8.89	87.7	
O4567-11	SE-SOIL-6	89	1.18	8.59	9.77	8.7	87.5	
O4567-12	SE-SOIL-6	90	1.19	8.63	9.82	8.71	87.1	
O4569-01	ETGI-367	91	1.15	8.79	9.94	8.47	83.3	
O4569-02	ETGI-367	92	1.12	8.87	9.99	9.00	88.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$