



Analytical Summary Report

Analysis Method: 1010B
Parameter: Flash Point
Run Number: LB127412
Thermometer ID: Flashpoint

Reviewed By: Rubina
Supervisor Review By: Iwona
Ambient Barometric Pressure (mmHg): 771.00
Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	09/21/2023	11:15
2	O4511-01		1	14	100.00	>212.0	>212.0	09/21/2023	11:45
3	O4511-02		1	12	100.00	>212.0	>212.0	09/21/2023	12:15
4	O4522-01		1	13	100.00	>212.0	>212.0	09/21/2023	12:45
5	O4537-02		1	14	100.00	>212.0	>212.0	09/21/2023	13:15
6	O4537-04		1	12	100.00	>212.0	>212.0	09/21/2023	13:45
7	O4540-01		1	11	100.00	>212.0	>212.0	09/21/2023	14:15
8	O4540-02		1	12	100.00	>212.0	>212.0	09/21/2023	14:45
9	O4540-03		1	14	100.00	>212.0	>212.0	09/21/2023	15:15
10	O4540-03DUP		1	14	100.00	>212.0	>212.0	09/21/2023	15:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16127412-

WorkList Name : fp-9-20

WorkList ID : 173757

Department : Wet-Chemistry

Date : 09-20-2023 08:06:35

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4511-01 E	50093	Water	Flash Point	Cool 4 deg C	PSEG03	D11	09/19/2023	1010B
O4511-02 E	50095	Water	Flash Point	Cool 4 deg C	PSEG03	D11	09/19/2023	1010B
O4522-01 C	33927	Water	Flash Point	Cool 4 deg C	PSEG03	D11	09/20/2023	1010B
O4537-02 C	COMP-2	Water	Flash Point	Cool 4 deg C	PSEG03	D21	09/20/2023	1010B
O4537-04 C	5866	Water	Flash Point	Cool 4 deg C	PSEG03	D21	09/20/2023	1010B
O4540-01 K	CENTRAL-M-R-HEATER	Water	Flash Point	Cool 4 deg C	PSEG03	N11	09/20/2023	1010B
O4540-02 Z	EMBRIDGE-M-R-HEATER	Water	Flash Point	Cool 4 deg C	PSEG03	N11	09/20/2023	1010B
O4540-03 X	TRANSCO-M-R-HEATER	Water	Flash Point	Cool 4 deg C	PSEG03	N11	09/20/2023	1010B

Date/Time 09/21/2023 11:00
 Raw Sample Received by: RDA CWO
 Raw Sample Relinquished by: JH GDCI

Date/Time 09/21/2023 15:50
 Raw Sample Received by: JH GDCI
 Raw Sample Relinquished by: RDA CWO