



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 9/25/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 13:50
In Date: 09/22/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:30
Out Date: 09/23/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB127417

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O4518-01	MJUB81	1	1.12	8.76	9.88	9.7	97.9	
O4518-02	MJUB82	2	1.13	8.66	9.79	9.06	91.6	
O4518-03	MJUB82D	3	1.13	8.66	9.79	9.06	91.6	
O4518-04	MJUB82S	4	1.13	8.66	9.79	9.06	91.6	
O4518-05	MJUB83	5	1.15	8.69	9.84	8.19	81.0	
O4518-06	MJUB84	6	1.14	8.81	9.95	9.02	89.4	
O4518-07	MJUB85	7	1.14	8.74	9.88	8.25	81.4	
O4518-08	MJUB86	8	1.18	8.59	9.77	8.5	85.2	
O4518-09	MJUB87	9	1.19	8.61	9.8	8.51	85.0	
O4518-10	MJUB88	10	1.13	8.69	9.82	8.13	80.6	
O4518-11	MJUB89	11	1.19	8.50	9.69	8.21	82.6	
O4518-12	MJUB90	12	1.15	8.53	9.68	8.3	83.8	
O4518-13	MJUB91	13	1.15	8.40	9.55	7.26	72.7	
O4518-14	MJUB92	14	1.19	8.42	9.61	6.81	66.7	
O4518-15	MJUB93	15	1.17	8.41	9.58	6.46	62.9	
O4518-16	MJUB94	16	1.1	8.79	9.89	6.91	66.1	
O4518-17	MJUB95	17	1.19	8.73	9.92	7.03	66.9	
O4518-18	MJUB96	18	1.19	8.38	9.57	7.06	70.0	
O4518-19	MJUB97	19	1.15	8.40	9.55	6.55	64.3	
O4518-20	MJUB98	20	1.14	8.82	9.96	6.84	64.6	
O4518-21	MJUB99	21	1.12	8.67	9.79	8.69	87.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$