

LB127425


245.1

INSTRUMENT ID: CV2

Linear

μ Abs.:

42158



A= 0.0000e+000

B= 2.3565e-004

C= -8.5098e-002

Rho= 0.9996664

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/1)
0.0	0.000	-0.006	-0.006	337	0.000	337					14
0.05	0.050	0.057	0.007	603	0.0 %	603	0				-7
0.2	0.200	0.186	-0.014	1151	0.0 %	1151					-2
2.5	2.500	2.443	-0.057	10727	0.0 %	10727					1
5.0	5.000	5.026	0.026	21691	0.0 %	21691					3
7.5	7.500	7.694	0.194	33009	0.0 %	33009					-2
10.0	10.000	9.850	-0.150	42158	0.0 %	42158					

LB127425 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 22 Sep 2023 11:28:02

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	337	-	0.000	245.1	PPB	22 Sep 2023 11:42:31	S	Std
0.05 - 1	50.05	603	-	0.050	245.1	PPB	22 Sep 2023 11:44:48	S	Std
0.2 - 1	50.2	1151	-	0.200	245.1	PPB	22 Sep 2023 11:49:53	S	Std
2.5 - 1	52.5	10727	-	2.500	245.1	PPB	22 Sep 2023 11:52:09	S	Std
5.0 - 1	55	21691	-	5.000	245.1	PPB	22 Sep 2023 11:54:25	S	Std
7.5 - 1	57.5	33009	-	7.500	245.1	PPB	22 Sep 2023 11:56:40	S	Std
10.0 - 1	57.0	42158	-	10.000	245.1	PPB	22 Sep 2023 12:04:32	S	Std
ICV76 - 1	ICV76	17113	3.9477	-	245.1	PPB	22 Sep 2023 12:07:21	U	SMPL
ICB76 - 1	ICB76	246	-0.0271	-	245.1	PPB	22 Sep 2023 12:09:36	U	SMPL
CCV43 - 1	CCV43	22121	5.1278	-	245.1	PPB	22 Sep 2023 12:11:51	U	SMPL
CCB43 - 1	CCB43	249	-0.0264	-	245.1	PPB	22 Sep 2023 12:14:06	U	SMPL
CRA - 1	CRA	561	0.0471	-	245.1	PPB	22 Sep 2023 12:16:22	U	SMPL
HighStd - 1	HighStd	42145	9.8466	-	245.1	PPB	22 Sep 2023 12:18:36	U	SMPL
ChkStd - 1	ChkStd	30753	7.1620	-	245.1	PPB	22 Sep 2023 12:20:51	U	SMPL
PB155750BL - 1	PBW	270	-0.0215	-	245.1	PPB	22 Sep 2023 12:25:53	U	SMPL
PB155750BS - 1	LCSW	17296	3.9908	-	245.1	PPB	22 Sep 2023 12:28:09	U	SMPL
O3572-10 - 1	LOD-MDL-WATER-04-QT3-2023	524	0.0384	-	245.1	PPB	22 Sep 2023 12:35:18	U	SMPL
O3572-10RE - 1	LOD-MDL-WATER-04-QT3-2023RE	927	0.1334	-	245.1	PPB	22 Sep 2023 12:37:34	U	SMPL
O3572-12 - 1	MDL-WATER-06-QT3-2023	583	0.0523	-	245.1	PPB	22 Sep 2023 12:42:26	U	SMPL
O4543-01 - 1	EFF-DAY-1-METALS	3216	0.6728	-	245.1	PPB	22 Sep 2023 12:50:06	U	SMPL
O4543-01DUP - 1	EFF-DAY-1-METALSDUP	3016	0.6256	-	245.1	PPB	22 Sep 2023 12:52:22	U	SMPL
CCV44 - 1	CCV44	21188	4.9080	-	245.1	PPB	22 Sep 2023 12:54:37	U	SMPL
CCB44 - 1	CCB44	237	-0.0292	-	245.1	PPB	22 Sep 2023 12:56:52	U	SMPL
O4543-01MS - 1	EFF-DAY-1-METALSMS	18237	4.2125	-	245.1	PPB	22 Sep 2023 12:59:07	U	SMPL
O4543-01MSD - 1	EFF-DAY-1-METALMSD	17845	4.1202	-	245.1	PPB	22 Sep 2023 13:01:22	U	SMPL
O4543-01LX5 - 1		807	0.1051	-	245.1	PPB	22 Sep 2023 13:03:39	U	SMPL
O4543-01A - 1		19137	4.4246	-	245.1	PPB	22 Sep 2023 13:05:56	U	SMPL
CCV45 - 1	CCV45	21510	4.9838	-	245.1	PPB	22 Sep 2023 13:08:11	U	SMPL
CCB45 - 1	CCB45	48	-0.0964	-	245.1	PPB	22 Sep 2023 13:10:28	U	SMPL