



TOTAL VOLATILE SOLIDS SM2540 G

TEMP1 IN: 104 °C 09/21/2023 16:40 TEMP1 OUT: 103 °C 09/22/2023 07:00
TEMP2 IN: 104 °C 09/22/2023 07:30 TEMP2 OUT: 104 °C 09/22/2023 09:00
TEMP3 IN: 540 °C 09/22/2023 09:30 TEMP3 OUT: 540 °C 09/22/2023 11:00
TEMP4 IN: 560 °C 09/22/2023 11:30 TEMP4 OUT: 560 °C 09/22/2023 13:00

Run Number: LB127427

SUPERVISOR: Iwona

ANALYST: jignesh

BalanceID: WC SC-6

OvenID: WC OVEN-1

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (mg/L)
1	LB127427BL	96.3871	96.3871	100	96.3871	96.3871	96.3871	96.3871	0.0000	0
2	O4574-01	89.1382	89.1382	100	89.5005	89.5005	89.4521	89.4521	0.0484	484
3	O4574-01DUP	92.6393	92.6393	100	93.0125	93.0125	92.9652	92.9652	0.0473	473

A = Sample Volume (ml)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105 °C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105 °C (g)

Weight D = C - B

Result % = $\frac{D}{A} \star 1000000$

VB 127427

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tvs water vo4574 WorkList ID : 173843 Department : Wet-Chemistry Date : 09-21-2023 16:17:17

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4574-01	FRAC-TANK-A2421	Water	TVS	Cool 4 deg C	PSEG03	N21	09/21/2023	SM2540 G

Date/Time 09.21.23 16:20
Raw Sample Received by: 10 (COC)
Raw Sample Relinquished by: CP SM

Date/Time 09.21.23 17:30
Raw Sample Received by: CP SM
Raw Sample Relinquished by: 20 (COC)