



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB127431

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	09/22/2023	10:20
2	O4563-02		1	14	100.00	>212.0	>212.0	09/22/2023	10:50
3	O4563-02DUP		1	12	100.00	>212.0	>212.0	09/22/2023	11:20
4	O4563-03		1	14	100.00	>212.0	>212.0	09/22/2023	11:50
5	O4574-01		1	14	100.00	>212.0	>212.0	09/22/2023	12:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6127431

WorkList Name : fp-9-22

WorkList ID : 173855

Department : Wet-Chemistry

Date : 09-22-2023 08:09:56

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4563-02 F	BEL-23-0006	Water	Flash Point	Cool 4 deg C	PSEG03	N11	09/21/2023	1010B
O4563-03 A	BEL-23-0008	Water	Flash Point	Cool 4 deg C	PSEG03	N11	09/21/2023	1010B
O4574-01 P	FRAC-TANK-A2421	Water	Flash Point	Cool 4 deg C	PSEG03	N21	09/21/2023	1010B

Date/Time 09/22/2023 10:15
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 09/22/2023 12:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]