



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB127432

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	09/22/2023	14:50
2	O4537-01		1	4	8.00	46.4	45.7	09/22/2023	15:20
3	O4537-01DUP		1	4	8.00	46.4	45.7	09/22/2023	15:50
4	O4563-06		1	11	61.00	141.8	141.1	09/22/2023	16:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6627432

WorkList Name : fp s 9-21

WorkList ID : 173818

Department : Wet-Chemistry

Date : 09-21-2023 10:53:04

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4537-01	COMP-1	Solid	Flash Point	Cool 4 deg C	PSEG03	D21	09/20/2023	1010B
O4563-06	BEL-23-0005	Solid	Flash Point	Cool 4 deg C	PSEG03	N11	09/21/2023	1010B

Date/Time 09/27/2023 14:45
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: JLC

Date/Time 09/22/2023 16:30
 Raw Sample Received by: JLC
 Raw Sample Relinquished by: RM CWC