



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB127444

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	09/22/2023	13:35
2	CAL2	1	Water	NA	NA	20.5	7.00	09/22/2023	13:36
3	CAL3	1	Water	NA	NA	20.6	10.02	09/22/2023	13:37
4	ICV	1	Water	NA	NA	20.5	7.01	09/22/2023	13:40
5	CCV1	1	Water	NA	NA	20.6	2.01	09/22/2023	13:45
6	O4563-02	1	Water	NA	NA	20.7	7.73	09/22/2023	14:20
7	O4563-02DUP	1	Water	NA	NA	20.8	7.74	09/22/2023	14:21
8	O4563-03	1	Water	NA	NA	20.4	6.02	09/22/2023	14:25
9	CCV2	1	Water	NA	NA	20.7	12.02	09/22/2023	14:32

WORKLIST(Hardcopy Internal Chain)

127444

WorkList Name : CORROSIVITY 04563 WorkList ID : 173881 Department : Wet-Chemistry Date : 09-22-2023 13:28:16

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4563-02	BEL-23-0006	Water	Corrosivity	Cool 4 deg C	PSEG03	N11	09/21/2023	9040C
O4563-03	BEL-23-0008	Water	Corrosivity	Cool 4 deg C	PSEG03	N11	09/21/2023	9040C

Date/Time 09-22-23 13:30

Raw Sample Received by: JAC

Raw Sample Relinquished by: JAC

Date/Time 09-22-23 17:30

Raw Sample Received by: JAC

Raw Sample Relinquished by: JAC